

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101222\
 Data File : VX032094.D
 Acq On : 12 Oct 2022 19:01
 Operator : JC/MD
 Sample : N5128-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 265751-4

Quant Time: Oct 13 09:22:43 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 06 14:28:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	86393	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	129317	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	135123	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	83095	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	93378	52.336	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	104.680%
35) Dibromofluoromethane	5.385	113	68643	46.894	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.780%
50) Toluene-d8	8.653	98	260809	49.515	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	99.040%
62) 4-Bromofluorobenzene	11.079	95	91979	49.395	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	98.780%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101222\
Data File : VX032094.D
Acq On : 12 Oct 2022 19:01
Operator : JC/MD
Sample : N5128-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
265751-4

Quant Time: Oct 13 09:22:43 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 06 14:28:46 2022
Response via : Initial Calibration

