

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101223\
 Data File : VX038246.D
 Acq On : 12 Oct 2023 11:01
 Operator : JC/MD
 Sample : VX1012WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX1012WBL01

Quant Time: Oct 13 04:44:22 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100523W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 05 15:27:13 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	114866	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	203820	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	204814	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	103633	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	84823	45.911	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	91.820%
35) Dibromofluoromethane	5.385	113	66329	45.163	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	90.320%
50) Toluene-d8	8.647	98	254031	45.265	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	90.540%#
62) 4-Bromofluorobenzene	11.079	95	96288	45.257	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	90.520%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101223\
Data File : VX038246.D
Acq On : 12 Oct 2023 11:01
Operator : JC/MD
Sample : VX1012WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX1012WBL01

Quant Time: Oct 13 04:44:22 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100523W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 05 15:27:13 2023
Response via : Initial Calibration

