

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101223\
 Data File : VX038249.D
 Acq On : 12 Oct 2023 12:15
 Operator : JC/MD
 Sample : 04859-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SVS-GROUNDWATER

Quant Time: Oct 13 04:46:13 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100523W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 05 15:27:13 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	110925	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	195001	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	197253	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	97556	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	80755	45.262	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	90.520%
35) Dibromofluoromethane	5.391	113	64915	46.199	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.400%
50) Toluene-d8	8.647	98	246116	45.838	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	91.680%#
62) 4-Bromofluorobenzene	11.079	95	91567	44.984	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	89.960%
Target Compounds						
16) Acetone	2.380	43	4280	5.748	ug/l	Qvalue # 88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101223\
Data File : VX038249.D
Acq On : 12 Oct 2023 12:15
Operator : JC/MD
Sample : 04859-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SVS-GROUNDWATER

Quant Time: Oct 13 04:46:13 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100523W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 05 15:27:13 2023
Response via : Initial Calibration

