

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101223\
 Data File : VX038260.D
 Acq On : 12 Oct 2023 16:36
 Operator : JC/MD
 Sample : 04875-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 GW-12-37-40

Quant Time: Oct 13 04:52:25 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100523W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 05 15:27:13 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	113259	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	202729	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	206896	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	104878	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	84530	46.402	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	92.800%
35) Dibromofluoromethane	5.385	113	65733	44.998	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	90.000%
50) Toluene-d8	8.647	98	258811	46.365	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	92.720%
62) 4-Bromofluorobenzene	11.079	95	96873	45.777	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	91.560%
Target Compounds						
					Qvalue	
16) Acetone	2.379	43	1589	2.090	ug/l #	89
44) Trichloroethene	7.129	130	17724	12.027	ug/l	92
64) Tetrachloroethene	9.275	164	21443	16.010	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101223\
Data File : VX038260.D
Acq On : 12 Oct 2023 16:36
Operator : JC/MD
Sample : 04875-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
GW-12-37-40

Quant Time: Oct 13 04:52:25 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100523W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 05 15:27:13 2023
Response via : Initial Calibration

