

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101223\
 Data File : VX038266.D
 Acq On : 12 Oct 2023 18:55
 Operator : JC/MD
 Sample : 04875-12
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 GW-14-37-40

Quant Time: Oct 13 04:54:26 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100523W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 05 15:27:13 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	110257	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	197716	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	206180	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	104092	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	83273	46.956	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	93.920%
35) Dibromofluoromethane	5.385	113	66076	46.380	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.760%
50) Toluene-d8	8.647	98	255011	46.842	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	93.680%
62) 4-Bromofluorobenzene	11.079	95	95763	46.399	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	92.800%
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	1507	2.036	ug/l #	88
44) Trichloroethene	7.129	130	3008	2.093	ug/l	95
64) Tetrachloroethene	9.275	164	3497	2.620	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101223\
Data File : VX038266.D
Acq On : 12 Oct 2023 18:55
Operator : JC/MD
Sample : 04875-12
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
GW-14-37-40

Quant Time: Oct 13 04:54:26 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100523W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 05 15:27:13 2023
Response via : Initial Calibration

