

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\WX101223\
 Data File : VX038244.D
 Acq On : 12 Oct 2023 10:04
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Oct 13 04:43:03 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100523W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 05 15:27:13 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	89	-0.01
2 T	Dichlorodifluoromethane	0.483	0.479	0.8	86	0.00
3 P	Chloromethane	0.492	0.451	8.3	88	0.00
4 C	Vinyl Chloride	0.545	0.516	5.3#	84	0.00
5 T	Bromomethane	0.744	0.742	0.3	86	0.00
6 T	Chloroethane	0.484	0.405	16.3	83	0.00
7 T	Trichlorofluoromethane	1.077	1.059	1.7	87	0.00
8 T	Diethyl Ether	0.402	0.390	3.0	87	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.470	0.478	-1.7	91	0.00
10 T	Methyl Iodide	0.677	0.627	7.4	83	0.00
11 T	Tert butyl alcohol	0.196	0.161	17.9	82	0.00
12 CM	1,1-Dichloroethene	0.457	0.442	3.3#	86	0.00
13 T	Acrolein	0.114	0.136	-19.3	118	0.00
14 T	Allyl chloride	0.818	0.760	7.1	83	0.00
15 T	Acrylonitrile	0.366	0.365	0.3	89	0.00
16 T	Acetone	0.336	0.401	-19.3	111	0.00
17 T	Carbon Disulfide	1.318	1.081	18.0	77	0.00
18 T	Methyl Acetate	1.439	1.444	-0.3	90	0.00
19 T	Methyl tert-butyl Ether	1.822	1.784	2.1	87	0.00
20 T	Methylene Chloride	0.631	0.561	11.1	87	0.00
21 T	trans-1,2-Dichloroethene	0.566	0.527	6.9	82	0.00
22 T	Diisopropyl ether	1.715	1.667	2.8	84	0.00
23 T	Vinyl Acetate	1.258	1.261	-0.2	86	0.00
24 P	1,1-Dichloroethane	1.041	0.970	6.8	84	0.00
25 T	2-Butanone	0.542	0.551	-1.7	91	0.00
26 T	2,2-Dichloropropane	0.813	0.776	4.6	83	0.00
27 T	cis-1,2-Dichloroethene	0.675	0.642	4.9	86	0.00
28 T	Bromochloromethane	0.470	0.489	-4.0	88	0.00
29 T	Tetrahydrofuran	0.353	0.337	4.5	86	0.00
30 C	Chloroform	1.138	1.129	0.8#	86	0.00
31 T	Cyclohexane	0.871	0.796	8.6	79	0.00
32 T	1,1,1-Trichloroethane	0.975	0.939	3.7	82	0.00
33 S	1,2-Dichloroethane-d4	0.804	0.787	2.1	86	-0.01
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	84	0.00
35 S	Dibromofluoromethane	0.360	0.370	-2.8	86	0.00
36 T	1,1-Dichloropropene	0.497	0.465	6.4	81	0.00
37 T	Ethyl Acetate	0.590	0.605	-2.5	88	0.00
38 T	Carbon Tetrachloride	0.489	0.490	-0.2	83	0.00
39 T	Methylcyclohexane	0.582	0.570	2.1	81	0.00
40 TM	Benzene	1.390	1.365	1.8	82	0.00
41 T	Methacrylonitrile	0.294	0.303	-3.1	87	-0.01
42 TM	1,2-Dichloroethane	0.555	0.557	-0.4	84	0.00
43 T	Isopropyl Acetate	0.881	0.902	-2.4	84	0.00
44 TM	Trichloroethene	0.363	0.370	-1.9	85	0.00
45 C	1,2-Dichloropropane	0.360	0.364	-1.1#	85	0.00
46 T	Dibromomethane	0.279	0.286	-2.5	87	0.00
47 T	Bromodichloromethane	0.496	0.534	-7.7	86	0.00
48 T	Methyl methacrylate	0.420	0.439	-4.5	85	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101223\
 Data File : VX038244.D
 Acq On : 12 Oct 2023 10:04
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Oct 13 04:43:03 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100523W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 05 15:27:13 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.010	0.011	-10.0	84	0.00
50 S	Toluene-d8	1.377	1.393	-1.2	85	0.00
51 T	4-Methyl-2-Pentanone	0.608	0.647	-6.4	89	0.00
52 CM	Toluene	0.905	0.921	-1.8#	84	0.00
53 T	t-1,3-Dichloropropene	0.533	0.582	-9.2	85	0.00
54 T	cis-1,3-Dichloropropene	0.560	0.604	-7.9	87	0.00
55 T	1,1,2-Trichloroethane	0.387	0.414	-7.0	89	0.00
56 T	Ethyl methacrylate	0.594	0.638	-7.4	87	0.00
57 T	1,3-Dichloropropane	0.626	0.664	-6.1	87	0.00
58 T	2-Chloroethyl Vinyl ether	0.291	0.300	-3.1	86	0.00
59 T	2-Hexanone	0.488	0.535	-9.6	92	0.00
60 T	Dibromochloromethane	0.379	0.435	-14.8	91	0.00
61 T	1,2-Dibromoethane	0.419	0.455	-8.6	89	0.00
62 S	4-Bromofluorobenzene	0.522	0.549	-5.2	89	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	88	0.00
64 T	Tetrachloroethene	0.324	0.313	3.4	84	0.00
65 PM	Chlorobenzene	1.046	1.030	1.5	88	0.00
66 T	1,1,1,2-Tetrachloroethane	0.354	0.376	-6.2	89	0.00
67 C	Ethyl Benzene	1.868	1.844	1.3#	86	0.00
68 T	m/p-Xylenes	0.722	0.702	2.8	85	0.00
69 T	o-Xylene	0.705	0.692	1.8	87	0.00
70 T	Styrene	1.148	1.181	-2.9	88	0.00
71 P	Bromoform	0.292	0.322	-10.3	92	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	91	0.00
73 T	Isopropylbenzene	3.353	3.256	2.9	87	0.00
74 T	N-ethyl acetate	1.432	1.467	-2.4	87	0.00
75 P	1,1,2,2-Tetrachloroethane	1.230	1.284	-4.4	94	0.00
76 T	1,2,3-Trichloropropane	1.137	1.202	-5.7	95	0.00
77 T	Bromobenzene	0.819	0.807	1.5	90	0.00
78 T	n-propylbenzene	4.100	4.015	2.1	88	0.00
79 T	2-Chlorotoluene	2.432	2.315	4.8	88	0.00
80 T	1,3,5-Trimethylbenzene	2.893	2.869	0.8	88	0.00
81 T	trans-1,4-Dichloro-2-butene	0.356	0.364	-2.2	88	0.00
82 T	4-Chlorotoluene	2.877	2.781	3.3	87	0.00
83 T	tert-Butylbenzene	2.827	2.815	0.4	89	0.00
84 T	1,2,4-Trimethylbenzene	2.884	2.890	-0.2	88	0.00
85 T	sec-Butylbenzene	3.662	3.662	0.0	88	0.00
86 T	p-Isopropyltoluene	3.021	3.053	-1.1	89	0.00
87 T	1,3-Dichlorobenzene	1.644	1.600	2.7	89	0.00
88 T	1,4-Dichlorobenzene	1.679	1.622	3.4	91	0.00
89 T	n-Butylbenzene	3.009	3.011	-0.1	89	0.00
90 T	Hexachloroethane	0.501	0.521	-4.0	88	0.00
91 T	1,2-Dichlorobenzene	1.581	1.580	0.1	92	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.297	0.314	-5.7	91	0.00
93 T	1,2,4-Trichlorobenzene	1.060	1.067	-0.7	93	0.00
94 T	Hexachlorobutadiene	0.427	0.447	-4.7	97	0.00
95 T	Naphthalene	3.692	3.642	1.4	90	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101223\
Data File : VX038244.D
Acq On : 12 Oct 2023 10:04
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Oct 13 04:43:03 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100523W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 05 15:27:13 2023
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	1.026	1.036	-1.0	92	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6