

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX101320\
 Data File : VX018926.D
 Acq On : 13 Oct 2020 12:37
 Operator : JC/SP
 Sample : VX1013MBL01
 Misc : 5.00µ/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX1013MBL01

Quant Time: Oct 14 02:11:21 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X100920W.M
 Quant Title : SW846 8260
 QLast Update : Sat Oct 10 06:09:09 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.62	168	260366	50.00	ug/l	-0.01
34) 1,4-Difluorobenzene	6.82	114	397004	50.00	ug/l	-0.01
63) Chlorobenzene-d5	10.10	117	366528	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	193506	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	184318	48.00	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.00%	
35) Dibromofluoromethane	5.46	113	141338	49.96	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.92%	
50) Toluene-d8	8.70	98	481878	49.10	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.20%	
62) 4-Bromofluorobenzene	11.12	95	164836	42.14	ug/l	0.00
Spiked Amount	50.000		Recovery	=	84.28%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX101320\
Data File : VX018926.D
Acq On : 13 Oct 2020 12:37
Operator : JC/SP
Sample : VX1013MBL01
Misc : 5.0µ/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX1013MBL01

Quant Time: Oct 14 02:11:21 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X100920W.M
Quant Title : SW846 8260
QLast Update : Sat Oct 10 06:09:09 2020
Response via : Initial Calibration

