

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101321\
 Data File : VX024800.D
 Acq On : 13 Oct 2021 21:56
 Operator : JC/MD
 Sample : M4190-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 P287-DITCH

Quant Time: Oct 14 03:58:50 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X101321W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 14 03:51:38 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	125471	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	220195	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	214985	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	88147	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	96495	62.57	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 125.14%#	
35) Dibromofluoromethane	5.397	113	76403	58.28	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 116.56%	
50) Toluene-d8	8.653	98	273619	56.43	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 112.86%#	
62) 4-Bromofluorobenzene	11.085	95	98051	51.41	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 102.82%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101321\
Data File : VX024800.D
Acq On : 13 Oct 2021 21:56
Operator : JC/MD
Sample : M4190-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
P287-DITCH

Quant Time: Oct 14 03:58:50 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X101321W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 14 03:51:38 2021
Response via : Initial Calibration

