

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101321\
 Data File : VX024803.D
 Acq On : 13 Oct 2021 23:06
 Operator : JC/MD
 Sample : M4190-07
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TRIP-BLANK

Quant Time: Oct 14 03:59:15 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X101321W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 14 03:51:38 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	127658	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	221834	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	213629	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	86573	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	98440	62.74	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 125.48%#	
35) Dibromofluoromethane	5.397	113	77066	58.36	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 116.72%	
50) Toluene-d8	8.653	98	273900	56.07	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 112.14%#	
62) 4-Bromofluorobenzene	11.085	95	97845	50.93	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 101.86%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101321\
Data File : VX024803.D
Acq On : 13 Oct 2021 23:06
Operator : JC/MD
Sample : M4190-07
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 29 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TRIP-BLANK

Quant Time: Oct 14 03:59:15 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X101321W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 14 03:51:38 2021
Response via : Initial Calibration

