

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101321\
 Data File : VX024796.D
 Acq On : 13 Oct 2021 20:21
 Operator : JC/MD
 Sample : M4170-01RE
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 P287-DITCHRE

Quant Time: Oct 14 03:58:16 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X101321W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 14 03:51:38 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	130045	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	230694	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	225141	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	99377	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	100505	62.880	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 125.760%#	
35) Dibromofluoromethane	5.397	113	79586	57.950	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 115.900%	
50) Toluene-d8	8.653	98	284263	55.956	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 111.920%	
62) 4-Bromofluorobenzene	11.085	95	106919	53.512	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 107.020%	


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101321\
Data File : VX024796.D
Acq On : 13 Oct 2021 20:21
Operator : JC/MD
Sample : M4170-01RE
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
P287-DITCHRE

Quant Time: Oct 14 03:58:16 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X101321W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 14 03:51:38 2021
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.562 min Scan# 734
Delta R.T. -0.000 min
Lab File: VX024796.D
Acq: 13 Oct 2021 20:21




Instrument :
MSVOA_X
ClientSampleId :
P287-DITCHRE


Tgt Ion:168 Resp: 130045
Ion Ratio Lower Upper
168 100
99 59.6 46.4 69.6



#33
1,2-Dichloroethane-d4
Concen: 62.880 ug/l
RT: 5.964 min Scan# 800
Delta R.T. -0.000 min
Lab File: VX024796.D
Acq: 13 Oct 2021 20:21

Tgt Ion: 65 Resp: 100505
Ion Ratio Lower Upper
65 100
67 50.6 0.0 101.0









#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 6.769 min Scan# 932
Delta R.T. 0.000 min
Lab File: VX024796.D
Acq: 13 Oct 2021 20:21




Instrument :
MSVOA_X
ClientSampleId :
P287-DITCHRE


Tgt Ion:114 Resp: 230694
Ion Ratio Lower Upper
114 100
63 18.8 0.0 41.0
88 15.8 0.0 34.8



#35
Dibromofluoromethane
Concen: 57.950 ug/l
RT: 5.397 min Scan# 707
Delta R.T. 0.006 min
Lab File: VX024796.D
Acq: 13 Oct 2021 20:21

Tgt Ion:113 Resp: 79586
Ion Ratio Lower Upper
113 100
111 102.6 82.1 123.1
192 19.7 15.6 23.4









#50
Toluene-d8
Concen: 55.956 ug/l
RT: 8.653 min Scan# 1241
Delta R.T. -0.000 min
Lab File: VX024796.D
Acq: 13 Oct 2021 20:21




Instrument :
MSVOA_X
ClientSampleId :
P287-DITCHRE


Tgt Ion: 98 Resp: 284263
Ion Ratio Lower Upper
98 100
100 64.3 51.4 77.0



#62
4-Bromofluorobenzene
Concen: 53.512 ug/l
RT: 11.085 min Scan# 1640
Delta R.T. -0.000 min
Lab File: VX024796.D
Acq: 13 Oct 2021 20:21

Tgt Ion: 95 Resp: 106919
Ion Ratio Lower Upper
95 100
174 80.1 0.0 158.0
176 77.7 0.0 153.6









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.055 min Scan# 1471
Delta R.T. 0.000 min
Lab File: VX024796.D
Acq: 13 Oct 2021 20:21

Instrument :
MSVOA_X
ClientSampleId :
P287-DITCHRE

Tgt Ion:117 Resp: 225141
Ion Ratio Lower Upper
117 100
82 58.7 44.6 67.0
119 33.0 24.8 37.2



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.024 min Scan# 1794
Delta R.T. 0.000 min
Lab File: VX024796.D
Acq: 13 Oct 2021 20:21

Tgt Ion:152 Resp: 99377
Ion Ratio Lower Upper
152 100
115 60.5 43.3 129.8
150 155.8 0.0 351.2

