

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101321\
 Data File : VX024799.D
 Acq On : 13 Oct 2021 21:32
 Operator : JC/MD
 Sample : M4170-07RE
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TRIP-BLANKRE

Quant Time: Oct 14 03:58:42 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X101321W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 14 03:51:38 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	126886	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	227178	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	223440	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	96974	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	100232	64.270	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 128.540%#	
35) Dibromofluoromethane	5.397	113	78576	58.100	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 116.200%	
50) Toluene-d8	8.653	98	283606	56.691	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 113.380%#	
62) 4-Bromofluorobenzene	11.085	95	105959	53.853	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 107.700%	


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101321\
Data File : VX024799.D
Acq On : 13 Oct 2021 21:32
Operator : JC/MD
Sample : M4170-07RE
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TRIP-BLANKRE

Quant Time: Oct 14 03:58:42 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X101321W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 14 03:51:38 2021
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.562 min Scan# 734
Delta R.T. -0.000 min
Lab File: VX024799.D
Acq: 13 Oct 2021 21:32




Instrument :
MSVOA_X
ClientSampleId :
TRIP-BLANKRE


Tgt Ion:168 Resp: 126886
Ion Ratio Lower Upper
168 100
99 59.7 46.4 69.6



#33
1,2-Dichloroethane-d4
Concen: 64.270 ug/l
RT: 5.964 min Scan# 800
Delta R.T. -0.000 min
Lab File: VX024799.D
Acq: 13 Oct 2021 21:32

Tgt Ion: 65 Resp: 100232
Ion Ratio Lower Upper
65 100
67 50.2 0.0 101.0









#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 6.769 min Scan# 932
Delta R.T. -0.000 min
Lab File: VX024799.D
Acq: 13 Oct 2021 21:32




Instrument :
MSVOA_X
ClientSampleId :
TRIP-BLANKRE


Tgt Ion:114 Resp: 227178
Ion Ratio Lower Upper
114 100
63 21.2 0.0 41.0
88 15.6 0.0 34.8



#35
Dibromofluoromethane
Concen: 58.100 ug/l
RT: 5.397 min Scan# 707
Delta R.T. 0.006 min
Lab File: VX024799.D
Acq: 13 Oct 2021 21:32

Tgt Ion:113 Resp: 78576
Ion Ratio Lower Upper
113 100
111 103.2 82.1 123.1
192 19.3 15.6 23.4









#50
Toluene-d8
Concen: 56.691 ug/l
RT: 8.653 min Scan# 1241
Delta R.T. -0.000 min
Lab File: VX024799.D
Acq: 13 Oct 2021 21:32




Instrument :
MSVOA_X
ClientSampleId :
TRIP-BLANKRE


Tgt Ion: 98 Resp: 283606
Ion Ratio Lower Upper
98 100
100 63.8 51.4 77.0



#62
4-Bromofluorobenzene
Concen: 53.853 ug/l
RT: 11.085 min Scan# 1640
Delta R.T. -0.000 min
Lab File: VX024799.D
Acq: 13 Oct 2021 21:32

Tgt Ion: 95 Resp: 105959
Ion Ratio Lower Upper
95 100
174 80.2 0.0 158.0
176 76.8 0.0 153.6









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.055 min Scan# 1471
Delta R.T. -0.000 min
Lab File: VX024799.D
Acq: 13 Oct 2021 21:32

Instrument :
MSVOA_X
ClientSampleId :
TRIP-BLANKRE

Tgt Ion:117 Resp: 223440
Ion Ratio Lower Upper
117 100
82 56.9 44.6 67.0
119 32.3 24.8 37.2



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.024 min Scan# 1794
Delta R.T. 0.000 min
Lab File: VX024799.D
Acq: 13 Oct 2021 21:32

Tgt Ion:152 Resp: 96974
Ion Ratio Lower Upper
152 100
115 61.2 43.3 129.8
150 157.8 0.0 351.2

