

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101322\
 Data File : VX032100.D
 Acq On : 13 Oct 2022 10:51
 Operator : JC/MD
 Sample : VX1013MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX1013MBL01

Quant Time: Oct 14 07:49:19 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 06 14:28:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	88371	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	132598	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	136360	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	85053	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	96023	52.614	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	105.220%
35) Dibromofluoromethane	5.379	113	71578	47.689	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.380%
50) Toluene-d8	8.647	98	265097	49.084	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	98.160%
62) 4-Bromofluorobenzene	11.079	95	94170	49.320	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	98.640%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101322\
Data File : VX032100.D
Acq On : 13 Oct 2022 10:51
Operator : JC/MD
Sample : VX1013MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX1013MBL01

Quant Time: Oct 14 07:49:19 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 06 14:28:46 2022
Response via : Initial Calibration

