

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101322\
 Data File : VX032104.D
 Acq On : 13 Oct 2022 12:36
 Operator : JC/MD
 Sample : N5132-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 GWOW-BR-01-340-360-101222

Quant Time: Oct 14 07:50:26 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 06 14:28:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	87698	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	130546	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	134842	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	81383	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	96543	53.305	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 106.600%			
35) Dibromofluoromethane	5.391	113	71785	48.579	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 97.160%			
50) Toluene-d8	8.647	98	263610	49.576	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 99.160%			
62) 4-Bromofluorobenzene	11.079	95	93051	49.500	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 99.000%			
Target Compounds					Qvalue	
10) Methyl Iodide	2.465	142	576	3.124	ug/l #	86
16) Acetone	2.386	43	5381	8.742	ug/l	89
52) Toluene	8.720	92	265934	70.885	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101322\
Data File : VX032104.D
Acq On : 13 Oct 2022 12:36
Operator : JC/MD
Sample : N5132-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
GWOW-BR-01-340-360-101222

Quant Time: Oct 14 07:50:26 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 06 14:28:46 2022
Response via : Initial Calibration

