

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101322\
 Data File : VX032115.D
 Acq On : 13 Oct 2022 17:00
 Operator : JC/MD
 Sample : N5046-23
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 R-4

Quant Time: Oct 14 07:52:58 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 06 14:28:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	82604	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	123028	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	127123	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	62969	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	91094	53.398	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	106.800%
35) Dibromofluoromethane	5.385	113	66309	47.615	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.240%
50) Toluene-d8	8.647	98	249662	49.822	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	99.640%
62) 4-Bromofluorobenzene	11.079	95	65938	37.220	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	74.440%#
Target Compounds						
						Qvalue
16) Acetone	2.386	43	6340	10.935	ug/l	94
20) Methylene Chloride	2.788	84	26365	17.305	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101322\
Data File : VX032115.D
Acq On : 13 Oct 2022 17:00
Operator : JC/MD
Sample : N5046-23
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
R-4

Quant Time: Oct 14 07:52:58 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 06 14:28:46 2022
Response via : Initial Calibration

