

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101322\
 Data File : VX032118.D
 Acq On : 13 Oct 2022 18:33
 Operator : JC/MD
 Sample : N5133-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 HAS-WATER-1013

Quant Time: Oct 14 07:53:27 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 06 14:28:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	83849	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	124343	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	129368	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	77572	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	91409	52.787	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 105.580%			
35) Dibromofluoromethane	5.391	113	68541	48.697	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 97.400%			
50) Toluene-d8	8.653	98	251082	49.575	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 99.160%			
62) 4-Bromofluorobenzene	11.079	95	87317	48.767	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 97.540%			
Target Compounds					Qvalue	
16) Acetone	2.386	43	1519	2.581	ug/l #	65
20) Methylene Chloride	2.788	84	789	0.510	ug/l #	83

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101322\
Data File : VX032118.D
Acq On : 13 Oct 2022 18:33
Operator : JC/MD
Sample : N5133-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
HAS-WATER-1013

Quant Time: Oct 14 07:53:27 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 06 14:28:46 2022
Response via : Initial Calibration

