



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101322\
 Data File : VX032098.D
 Acq On : 13 Oct 2022 09:48
 Operator : JC/MD
 Sample : BFB
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100622W.M
 Title : SW846 8260
 Last Update : Thu Oct 06 14:28:46 2022



AutoFind: Scans 1638, 1639, 1640; Background Corrected with Scan 1632

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result
50	95	15	40	21.5	9713	PASS
75	95	30	60	56.7	25600	PASS
95	95	100	100	100.0	45187	PASS
96	95	5	9	6.3	2853	PASS
173	174	0.00	2	1.2	388	PASS
174	95	50	100	73.9	33376	PASS
175	174	5	9	7.3	2420	PASS
176	174	95	101	95.3	31816	PASS
177	176	5	9	6.7	2135	PASS