

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101325\
 Data File : VX048132.D
 Acq On : 13 Oct 2025 11:44
 Operator : JC/MD
 Sample : Q3324-05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB-184-GW-780-782

Quant Time: Oct 14 01:52:19 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 17 06:39:58 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	145285	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.751	114	293131	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	303093	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	152502	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	116568	50.406	ug/l	0.00
Spiked Amount	50.000	Range 78 - 117	Recovery	= 100.820%		
35) Dibromofluoromethane	5.373	113	96052	48.859	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 97.720%		
50) Toluene-d8	8.635	98	331940	48.798	ug/l	0.00
Spiked Amount	50.000	Range 92 - 112	Recovery	= 97.600%		
62) 4-Bromofluorobenzene	11.061	95	145147	56.223	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	= 112.440%		
Target Compounds						
16) Acetone	2.374	43	3821	3.800	ug/l	Qvalue 96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101325\
Data File : VX048132.D
Acq On : 13 Oct 2025 11:44
Operator : JC/MD
Sample : Q3324-05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-184-GW-780-782

Quant Time: Oct 14 01:52:19 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 17 06:39:58 2025
Response via : Initial Calibration

