

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101325\
 Data File : VX048138.D
 Acq On : 13 Oct 2025 13:47
 Operator : JC/MD
 Sample : Q3311-13DL 5X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-215DDL

Quant Time: Oct 14 01:54:24 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 17 06:39:58 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	135078	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.751	114	281533	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	297384	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	150578	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	117332	54.570	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 109.140%	
35) Dibromofluoromethane	5.379	113	94206	49.894	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 99.780%	
50) Toluene-d8	8.635	98	324424	49.657	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 99.320%	
62) 4-Bromofluorobenzene	11.061	95	139992	56.460	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 112.920%	
Target Compounds						
64) Tetrachloroethene	9.257	164	222499	114.771	ug/l	Qvalue 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101325\
Data File : VX048138.D
Acq On : 13 Oct 2025 13:47
Operator : JC/MD
Sample : Q3311-13DL 5X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-215DDL

Quant Time: Oct 14 01:54:24 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 17 06:39:58 2025
Response via : Initial Calibration

