

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101325\
 Data File : VX048139.D
 Acq On : 13 Oct 2025 14:08
 Operator : JC/MD
 Sample : Q3311-14 200X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-227B

Quant Time: Oct 14 01:54:44 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 17 06:39:58 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	132723	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.751	114	271642	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	287637	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	145860	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	114891	54.383	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 108.760%	
35) Dibromofluoromethane	5.379	113	92359	50.697	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 101.400%	
50) Toluene-d8	8.635	98	308796	48.986	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 97.980%	
62) 4-Bromofluorobenzene	11.067	95	137711	57.562	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 115.120%	
Target Compounds						
64) Tetrachloroethene	9.262	164	3791	2.022	ug/l	Qvalue 88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101325\
Data File : VX048139.D
Acq On : 13 Oct 2025 14:08
Operator : JC/MD
Sample : Q3311-14 200X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-227B

Quant Time: Oct 14 01:54:44 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 17 06:39:58 2025
Response via : Initial Calibration

