

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101325\
 Data File : VX048142.D
 Acq On : 13 Oct 2025 15:10
 Operator : JC/MD
 Sample : PB170063TB
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB170063TB

Quant Time: Oct 14 01:55:40 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 17 06:39:58 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.538	168	122716	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.745	114	253968	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	271431	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	136338	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.940	65	103817	53.149	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 106.300%	
35) Dibromofluoromethane	5.379	113	84014	49.326	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 98.660%	
50) Toluene-d8	8.635	98	293820	49.854	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 99.700%	
62) 4-Bromofluorobenzene	11.061	95	126841	56.709	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 113.420%	
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	27468	32.341	ug/l	98
18) Methyl Acetate	2.703	43	8960	4.740	ug/l	99
20) Methylene Chloride	2.794	84	10188	5.667	ug/l	95
25) 2-Butanone	4.556	43	7866	7.425	ug/l #	86
37) Ethyl Acetate	4.721	43	6172	2.316	ug/l #	89
43) Isopropyl Acetate	6.336	43	7949	1.819	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101325\
Data File : VX048142.D
Acq On : 13 Oct 2025 15:10
Operator : JC/MD
Sample : PB170063TB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB170063TB

Quant Time: Oct 14 01:55:40 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 17 06:39:58 2025
Response via : Initial Calibration

