

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101325\
 Data File : VX048143.D
 Acq On : 13 Oct 2025 15:30
 Operator : JC/MD
 Sample : Q3321-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 P001-DS-01

Quant Time: Oct 14 01:56:05 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 17 06:39:58 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	125413	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.751	114	262380	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	277983	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	140622	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.940	65	112725	56.468	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	112.940%
35) Dibromofluoromethane	5.379	113	90650	51.516	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	103.040%
50) Toluene-d8	8.634	98	308347	50.642	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	101.280%
62) 4-Bromofluorobenzene	11.067	95	132352	57.275	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	114.560%
Target Compounds						
						Qvalue
16) Acetone	2.379	43	37321	42.998	ug/l	99
18) Methyl Acetate	2.709	43	3801	1.968	ug/l	93
20) Methylene Chloride	2.794	84	5215	2.838	ug/l	93
43) Isopropyl Acetate	6.336	43	5035	1.115	ug/l #	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101325\
Data File : VX048143.D
Acq On : 13 Oct 2025 15:30
Operator : JC/MD
Sample : Q3321-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
P001-DS-01

Quant Time: Oct 14 01:56:05 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 17 06:39:58 2025
Response via : Initial Calibration

