

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101325\
 Data File : VX048145.D
 Acq On : 13 Oct 2025 16:11
 Operator : JC/MD
 Sample : Q3326-08
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TP-2

Quant Time: Oct 14 01:56:51 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 17 06:39:58 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	117555	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.751	114	241874	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	256421	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	127657	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	99821	53.347	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	106.700%
35) Dibromofluoromethane	5.373	113	81505	50.245	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	100.500%
50) Toluene-d8	8.635	98	285803	50.919	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	101.840%
62) 4-Bromofluorobenzene	11.061	95	120716	56.669	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	113.340%
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	34742	42.702	ug/l	100
18) Methyl Acetate	2.703	43	4580	2.529	ug/l #	87
20) Methylene Chloride	2.788	84	6556	3.807	ug/l	93
43) Isopropyl Acetate	6.330	43	11094	2.665	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101325\
Data File : VX048145.D
Acq On : 13 Oct 2025 16:11
Operator : JC/MD
Sample : Q3326-08
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TP-2

Quant Time: Oct 14 01:56:51 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 17 06:39:58 2025
Response via : Initial Calibration

