

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101325\
 Data File : VX048148.D
 Acq On : 13 Oct 2025 17:13
 Operator : JC/MD
 Sample : Q3311-02 200X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-219

Quant Time: Oct 14 01:57:59 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 17 06:39:58 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	116144	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.751	114	238626	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	251599	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	128285	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.940	65	96549	52.225	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	104.440%
35) Dibromofluoromethane	5.379	113	76588	47.857	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.720%
50) Toluene-d8	8.635	98	267434	48.295	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	96.580%
62) 4-Bromofluorobenzene	11.067	95	115786	55.094	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	110.180%
Target Compounds						
64) Tetrachloroethene	9.256	164	825	0.503	ug/l	Qvalue 89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101325\
Data File : VX048148.D
Acq On : 13 Oct 2025 17:13
Operator : JC/MD
Sample : Q3311-02 200X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-219

Quant Time: Oct 14 01:57:59 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 17 06:39:58 2025
Response via : Initial Calibration

