

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\WX101419\
 Data File : VX013065.D
 Acq On : 14 Oct 2019 20:22
 Operator : JC/SP
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :

Quant Time: Oct 15 01:30:45 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X100819W.M
 Quant Title : SW846 8260
 QLast Update : Fri Oct 11 09:48:22 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	143288	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	251722	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	231619	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	111222	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.05	65	102335	57.14	ug/l	0.00
Spiked Amount			Recovery	=	114.28%	
35) Dibromofluoromethane	5.49	113	80848	55.96	ug/l	0.00
Spiked Amount			Recovery	=	111.92%	
50) Toluene-d8	8.71	98	295793	52.49	ug/l	0.00
Spiked Amount			Recovery	=	104.98%	
62) 4-Bromofluorobenzene	11.14	95	111654	51.00	ug/l	0.00
Spiked Amount			Recovery	=	102.00%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
2) Dichlorodifluoromethane	1.19	85	63255	43.309	ug/l	98
3) Chloromethane	1.31	50	74492	46.864	ug/l	100
4) Vinyl Chloride	1.40	62	72565	44.229	ug/l	99
5) Bromomethane	1.62	94	30197	44.277	ug/l	97
6) Chloroethane	1.70	64	49242	49.142	ug/l	99
7) Trichlorofluoromethane	1.91	101	104008	46.969	ug/l	97
8) Diethyl Ether	2.18	74	48202	53.928	ug/l	100
9) 1,1,2-Trichlorotrifluoroet	2.37	101	67100	46.820	ug/l	99
10) Methyl Iodide	2.50	142	87718	53.617	ug/l	99
11) Tert butyl alcohol	3.04	59	127969	269.547	ug/l	99
12) 1,1-Dichloroethene	2.36	96	69471	47.977	ug/l	93
13) Acrolein	2.28	56	73409	217.580	ug/l	98
14) Allyl chloride	2.72	41	130479	50.599	ug/l	99
15) Acrylonitrile	3.13	53	267174	296.960	ug/l	99
16) Acetone	2.44	43	226900	240.362	ug/l	99
17) Carbon Disulfide	2.56	76	159248	38.570	ug/l	100
18) Methyl Acetate	2.76	43	131504	58.772	ug/l	99
19) Methyl tert-butyl Ether	3.19	73	287479	57.491	ug/l	98
20) Methylene Chloride	2.84	84	89462	51.567	ug/l	95
21) trans-1,2-Dichloroethene	3.16	96	77695	50.062	ug/l	95
22) Diisopropyl ether	3.84	45	273365	55.323	ug/l	93
23) Vinyl Acetate	3.80	43	1187912	278.484	ug/l	100
24) 1,1-Dichloroethane	3.69	63	149383	53.739	ug/l	99
25) 2-Butanone	4.66	43	365750	279.271	ug/l	97
26) 2,2-Dichloropropane	4.58	77	105827	45.272	ug/l	98
27) cis-1,2-Dichloroethene	4.58	96	95498	53.382	ug/l	97
28) Bromochloromethane	5.01	49	49143	59.237	ug/l	95
29) Tetrahydrofuran	5.12	42	233967	290.311	ug/l	97
30) Chloroform	5.20	83	153676	53.877	ug/l	99
31) Cyclohexane	5.57	56	114773	44.155	ug/l	97
32) 1,1,1-Trichloroethane	5.48	97	124505	51.517	ug/l	99
36) 1,1-Dichloropropene	5.79	75	103018	46.548	ug/l	99
37) Ethyl Acetate	4.82	43	138664	56.032	ug/l	99
38) Carbon Tetrachloride	5.77	117	102852	47.746	ug/l	98

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX101419\
 Data File : VX013065.D
 Acq On : 14 Oct 2019 20:22
 Operator : JC/SP
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :

Quant Time: Oct 15 01:30:45 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X100819W.M
 Quant Title : SW846 8260
 QLast Update : Fri Oct 11 09:48:22 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
39) Methylcyclohexane	7.46	83	116194	43.292	ug/l	98
40) Benzene	6.14	78	341396	52.229	ug/l	99
41) Methacrylonitrile	5.03	41	74746	54.967	ug/l	98
42) 1,2-Dichloroethane	6.19	62	126007	53.723	ug/l	98
43) Isopropyl Acetate	6.44	43	218357	54.974	ug/l	100
44) Trichloroethene	7.21	130	84790	46.778	ug/l	94
45) 1,2-Dichloropropane	7.51	63	89475	53.708	ug/l	98
46) Dibromomethane	7.65	93	60454	54.465	ug/l	97
47) Bromodichloromethane	7.89	83	116348	53.713	ug/l	99
48) Methyl methacrylate	7.76	41	104215	53.841	ug/l	96
49) 1,4-Dioxane	7.73	88	55690	1069.602	ug/l	99
51) 4-Methyl-2-Pentanone	8.64	43	701295	283.184	ug/l	99
52) Toluene	8.78	92	213848	51.138	ug/l	98
53) t-1,3-Dichloropropene	9.04	75	126991	47.109	ug/l	98
54) cis-1,3-Dichloropropene	8.43	75	140442	52.716	ug/l	99
55) 1,1,2-Trichloroethane	9.21	97	92464	55.872	ug/l	97
56) Ethyl methacrylate	9.17	69	151354	51.034	ug/l	98
57) 1,3-Dichloropropane	9.37	76	155567	54.342	ug/l	99
58) 2-Chloroethyl Vinyl ether	8.31	63	343053	282.175	ug/l	98
59) 2-Hexanone	9.49	43	546095	283.278	ug/l	100
60) Dibromochloromethane	9.58	129	91772	49.838	ug/l	100
61) 1,2-Dibromoethane	9.67	107	94189	54.104	ug/l	99
64) Tetrachloroethene	9.33	164	80285	49.975	ug/l	95
65) Chlorobenzene	10.14	112	227730	49.511	ug/l	99
66) 1,1,1,2-Tetrachloroethane	10.21	131	85089	53.355	ug/l	99
67) Ethyl Benzene	10.25	91	397141	48.609	ug/l	96
68) m/p-Xylenes	10.36	106	304554	99.781	ug/l	99
69) o-Xylene	10.70	106	149913	50.366	ug/l	99
70) Styrene	10.71	104	263399	52.055	ug/l	99
71) Bromoform	10.86	173	63741	47.110	ug/l #	99
73) Isopropylbenzene	11.01	105	380550	46.538	ug/l	99
74) N-amyl acetate	10.89	43	186030	52.823	ug/l	99
75) 1,1,2,2-Tetrachloroethane	11.26	83	141827	52.147	ug/l	99
76) 1,2,3-Trichloropropane	11.29	75	159094	64.492	ug/l	80
77) Bromobenzene	11.25	156	96431	50.043	ug/l	97
78) n-propylbenzene	11.35	91	422645	46.971	ug/l	99
79) 2-Chlorotoluene	11.42	91	266592	47.143	ug/l	99
80) 1,3,5-Trimethylbenzene	11.50	105	329880	48.313	ug/l	98
81) trans-1,4-Dichloro-2-buten	11.07	75	43880	46.088	ug/l	100
82) 4-Chlorotoluene	11.51	91	314731	48.557	ug/l	99
83) tert-Butylbenzene	11.76	119	315865	47.422	ug/l	99
84) 1,2,4-Trimethylbenzene	11.80	105	343039	49.606	ug/l	99
85) sec-Butylbenzene	11.94	105	351141	45.504	ug/l	98
86) p-Isopropyltoluene	12.06	119	326378	45.604	ug/l	99
87) 1,3-Dichlorobenzene	12.02	146	172471	48.198	ug/l	100
88) 1,4-Dichlorobenzene	12.09	146	174370	48.115	ug/l	98
89) n-Butylbenzene	12.39	91	260945	43.383	ug/l	97
90) Hexachloroethane	12.59	117	53490	40.762	ug/l	95
91) 1,2-Dichlorobenzene	12.39	146	175307	49.972	ug/l	99
92) 1,2-Dibromo-3-Chloropropan	12.99	75	35082	55.033	ug/l	99


Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX101419\
Data File : VX013065.D
Acq On : 14 Oct 2019 20:22
Operator : JC/SP
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 30 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :

Quant Time: Oct 15 01:30:45 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X100819W.M
Quant Title : SW846 8260
QLast Update : Fri Oct 11 09:48:22 2019
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
93) 1,2,4-Trichlorobenzene	13.64	180	105816	48.152	ug/l	99
94) Hexachlorobutadiene	13.78	225	41923	41.337	ug/l	98
95) Naphthalene	13.83	128	398331	52.289	ug/l	100
96) 1,2,3-Trichlorobenzene	14.01	180	106960	47.100	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed



#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 5.65 min Scan# 782

Delta R.T. 0.00 min

Lab File: VX013065.D

Acq: 14 Oct 2019 20:22

Instrument :

MSVOA_X


ClientSampleId :

Tot Ion:168 Resp: 143288

Ion Ratio Lower Upper

168 100

99 59.7 49.4 74.0



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VX0

5.65

60000

50000

40000



30000

20000

10000

0

Time-->



#2

Dichlorodifluoromethane

Concen: 43.309 ug/l

RT: 1.19 min Scan# 49

Delta R.T. -0.01 min

Lab File: VX013065.D


Acq: 14 Oct 2019 20:22

Tgt Ion: 85 Resp: 63255

Ion Ratio Lower Upper

85 100

87 34.0 16.4 49.4



Abundance Ion 84.90 (84.60 to 85.60): VX0

Ion 86.90 (86.60 to 87.60): VX0

1.19


60000


40000

20000

0

Time-->





#3

Chloromethane

Concen: 46.864 ug/l

RT: 1.31 min Scan# 70

Delta R.T. -0.01 min

Lab File: VX013065.D

Acq: 14 Oct 2019 20:22

Instrument :

MSVOA_X


ClientSampleId :

Tot Ion: 50 Resp: 74492

Ion Ratio Lower Upper




50 100

52 31.5 25.2 37.8



Abundance Ion 49.90 (49.60 to 50.60): VX0

Ion 51.90 (51.60 to 52.60): VX0



#4

Vinyl Chloride

Concen: 44.229 ug/l

RT: 1.40 min Scan# 84

Delta R.T. 0.00 min

Lab File: VX013065.D


Acq: 14 Oct 2019 20:22

Tgt Ion: 62 Resp: 72565

Ion Ratio Lower Upper



62 100


64 32.0 25.3 37.9



Abundance Ion 61.90 (61.60 to 62.60): VX0

Ion 63.90 (63.60 to 64.60): VX0





#5

Bromomethane

Concen: 44.277 ug/l

RT: 1.62 min Scan# 121

Delta R.T. -0.01 min

Lab File: VX013065.D

Acq: 14 Oct 2019 20:22

Instrument :

MSVOA_X


ClientSampleId :

Tot Ion: 94 Resp: 30197

Ion Ratio Lower Upper




94 100

96 90.8 74.7 112.1



Abundance Ion 93.90 (93.60 to 94.60): VX0

Ion 95.90 (95.60 to 96.60): VX0



#6

Chloroethane

Concen: 49.142 ug/l

RT: 1.70 min Scan# 134

Delta R.T. -0.01 min

Lab File: VX013065.D


Acq: 14 Oct 2019 20:22

Tgt Ion: 64 Resp: 49242

Ion Ratio Lower Upper



64 100


66 31.6 25.6 38.4



Abundance Ion 63.90 (63.60 to 64.60): VX0

Ion 65.90 (65.60 to 66.60): VX0









#7

Trichlorofluoromethane
 Concen: 46.969 ug/l
 RT: 1.91 min Scan# 168
 Delta R.T. -0.01 min
 Lab File: VX013065.D
 Acq: 14 Oct 2019 20:22

Instrument :
 MSVOA_X
 ClientSampleId :




Tot Ion:101 Resp: 104008
 Ion Ratio Lower Upper
 101 100
 103 63.3 52.5 78.7




#8

Diethyl Ether
 Concen: 53.928 ug/l
 RT: 2.18 min Scan# 212
 Delta R.T. 0.00 min
 Lab File: VX013065.D
 Acq: 14 Oct 2019 20:22

Tgt Ion: 74 Resp: 48202
 Ion Ratio Lower Upper
 74 100
 45 95.6 47.8 143.4





#9

1,1,2-Trichlorotrifluoroethane

Concen: 46.820 ug/l

RT: 2.37 min Scan# 243

Delta R.T. -0.01 min

Lab File: VX013065.D

Acq: 14 Oct 2019 20:22

Instrument :

MSVOA_X

ClientSampled :


Tot Ion:101 Resp: 67100

Ion Ratio Lower Upper

101 100

85 41.7 35.2 52.8




151 75.4 60.3 90.5



Abundance Ion 100.90 (100.60 to 101.60): V

Ion 84.90 (84.60 to 85.60): VX0

Ion 150.80 (150.50 to 151.50): V



#10

Methyl Iodide

Concen: 53.617 ug/l

RT: 2.50 min Scan# 264

Delta R.T. -0.01 min

Lab File: VX013065.D

Acq: 14 Oct 2019 20:22


Tgt Ion:142 Resp: 87718

Ion Ratio Lower Upper

142 100

127 44.0 34.9 52.3



141 14.3 11.6 17.4




Abundance Ion 141.80 (141.50 to 142.50): V

Ion 126.80 (126.50 to 127.50): V

Ion 140.80 (140.50 to 141.50): V









#11

Tert butyl alcohol
Concen: 269.547 ug/l
RT: 3.04 min Scan# 353
Delta R.T. 0.00 min
Lab File: VX013065.D
Acq: 14 Oct 2019 20:22

Instrument :
MSVOA_X
ClientSampleId :




Tot Ion: 59 Resp: 127969
Ion Ratio Lower Upper
59 100
57 10.0 8.3 12.5




#12

1,1-Dichloroethene
Concen: 47.977 ug/l
RT: 2.36 min Scan# 242
Delta R.T. 0.00 min
Lab File: VX013065.D
Acq: 14 Oct 2019 20:22

Tgt Ion: 96 Resp: 69471
Ion Ratio Lower Upper
96 100
61 154.0 132.2 198.4
98 65.3 50.5 75.7





#13

Acrolein

Concen: 217.580 ug/l

RT: 2.28 min Scan# 229

Delta R.T. 0.00 min

Lab File: VX013065.D

Acq: 14 Oct 2019 20:22





Instrument :
MSVOA_X
ClientSampled :

Tot Ion: 56 Resp: 73409

Ion Ratio Lower Upper

56 100

55 69.5 56.6 85.0



#14

Allyl chloride

Concen: 50.599 ug/l

RT: 2.72 min Scan# 300

Delta R.T. 0.00 min

Lab File: VX013065.D

Acq: 14 Oct 2019 20:22




Tgt Ion: 41 Resp: 130479


Ion Ratio Lower Upper

41 100

39 62.1 51.0 76.6

76 32.7 26.2 39.4





#15

Acrylonitrile

Concen: 296.960 ug/l

RT: 3.13 min Scan# 368

Delta R.T. -0.01 min

Lab File: VX013065.D

Acq: 14 Oct 2019 20:22

Instrument :

MSVOA_X

ClientSampleId :





Tot Ion: 53 Resp: 267174

Ion Ratio Lower Upper

53 100

52 81.8 66.2 99.2

51 36.1 28.2 42.4



#16

Acetone

Concen: 240.362 ug/l

RT: 2.44 min Scan# 254

Delta R.T. 0.00 min

Lab File: VX013065.D




Acq: 14 Oct 2019 20:22


Tgt Ion: 43 Resp: 226900

Ion Ratio Lower Upper

43 100

58 31.3 25.4 38.2





#17

Carbon Disulfide

Concen: 38.570 ug/l

RT: 2.56 min Scan# 274

Delta R.T. -0.01 min

Lab File: VX013065.D

Acq: 14 Oct 2019 20:22

Instrument :

MSVOA_X


ClientSampleId :

Tot Ion: 76 Resp: 159248

Ion Ratio Lower Upper




76 100

78 8.8 7.1 10.7



Abundance Ion 75.90 (75.60 to 76.60): VX0

Ion 77.90 (77.60 to 78.60): VX0



#18

Methyl Acetate

Concen: 58.772 ug/l

RT: 2.76 min Scan# 308

Delta R.T. 0.00 min

Lab File: VX013065.D


Acq: 14 Oct 2019 20:22

Tgt Ion: 43 Resp: 131504

Ion Ratio Lower Upper



43 100


74 24.6 19.2 28.8



Abundance Ion 43.00 (42.70 to 43.70): VX0

Ion 74.00 (73.70 to 74.70): VX0









#19
Methyl tert-butyl Ether
Concen: 57.491 ug/l
RT: 3.19 min Scan# 377
Delta R.T. 0.00 min
Lab File: VX013065.D
Acq: 14 Oct 2019 20:22




Instrument :
MSVOA_X
ClientSampled :


Tot Ion: 73 Resp: 287479
Ion Ratio Lower Upper
73 100
57 21.2 17.8 26.6



#20
Methylene Chloride
Concen: 51.567 ug/l
RT: 2.84 min Scan# 321
Delta R.T. -0.01 min
Lab File: VX013065.D
Acq: 14 Oct 2019 20:22

Tgt Ion: 84 Resp: 89462
Ion Ratio Lower Upper
84 100
49 112.3 96.6 144.8
51 37.1 30.7 46.1
86 64.4 51.0 76.6





#21

trans-1,2-Dichloroethene

Concen: 50.062 ug/l

RT: 3.16 min Scan# 372

Delta R.T. 0.00 min

Lab File: VX013065.D

Acq: 14 Oct 2019 20:22

Instrument :

MSVOA_X

ClientSampleId :


Tot Ion: 96 Resp: 77695

Ion Ratio Lower Upper

96 100

61 130.9 110.4 165.6




98 61.2 51.5 77.3



Abundance Ion 95.90 (95.60 to 96.60): VX0

Ion 60.90 (60.60 to 61.60): VX0

Ion 97.90 (97.60 to 98.60): VX0



#22

Diisopropyl ether

Concen: 55.323 ug/l

RT: 3.84 min Scan# 485

Delta R.T. -0.01 min

Lab File: VX013065.D

Acq: 14 Oct 2019 20:22

Tgt Ion: 45 Resp: 273365


Ion Ratio Lower Upper

45 100

43 72.2 51.4 77.0

87 29.7 22.5 33.7

59 11.4 9.7 14.5






Abundance Ion 45.00 (44.70 to 45.70): VX0

Ion 43.00 (42.70 to 43.70): VX0

Ion 87.00 (86.70 to 87.70): VX0

Ion 59.00 (58.70 to 59.70): VX0





#23

Vinyl Acetate

Concen: 278.484 ug/l

RT: 3.80 min Scan# 478

Delta R.T. -0.01 min

Lab File: VX013065.D

Acq: 14 Oct 2019 20:22

Instrument :

MSVOA_X


ClientSampled :

Tot Ion: 43 Resp: 1187912

Ion Ratio Lower Upper




43 100

86 11.2 9.0 13.6



Abundance Ion 43.00 (42.70 to 43.70): VX0

Ion 86.00 (85.70 to 86.70): VX0



#24

1,1-Dichloroethane

Concen: 53.739 ug/l

RT: 3.69 min Scan# 459

Delta R.T. 0.00 min

Lab File: VX013065.D

Acq: 14 Oct 2019 20:22


Tgt Ion: 63 Resp: 149383

Ion Ratio Lower Upper

63 100

98 7.8 3.9 11.5



100 5.2 2.3 6.9




Abundance Ion 62.90 (62.60 to 63.60): VX0

Ion 97.90 (97.60 to 98.60): VX0

Ion 99.90 (99.60 to 100.60): VX0





#25

2-Butanone

Concen: 279.271 ug/l

RT: 4.66 min Scan# 619

Delta R.T. 0.00 min

Lab File: VX013065.D

Acq: 14 Oct 2019 20:22





Instrument :
MSVOA_X
ClientSampled :

Tot Ion: 43 Resp: 365750

Ion Ratio Lower Upper

43 100

72 27.8 21.0 31.6



#26

2,2-Dichloropropane

Concen: 45.272 ug/l

RT: 4.58 min Scan# 605

Delta R.T. 0.00 min

Lab File: VX013065.D




Acq: 14 Oct 2019 20:22


Tgt Ion: 77 Resp: 105827

Ion Ratio Lower Upper

77 100

97 23.7 11.3 33.9






#27




cis-1,2-Dichloroethene
 Concen: 53.382 ug/l
 RT: 4.58 min Scan# 606
 Delta R.T. -0.01 min
 Lab File: VX013065.D
 Acq: 14 Oct 2019 20:22

Instrument :
 MSVOA_X
 ClientSampled :

Tot Ion: 96 Resp: 95498
 Ion Ratio Lower Upper
 96 100
 61 140.5 0.0 290.6
 98 64.7 0.0 128.8




Abundance Ion 95.90 (95.60 to 96.60): VX0
 Ion 60.90 (60.60 to 61.60): VX0
 Ion 97.90 (97.60 to 98.60): VX0





#28


Bromochloromethane
 Concen: 59.237 ug/l
 RT: 5.01 min Scan# 676
 Delta R.T. 0.00 min
 Lab File: VX013065.D
 Acq: 14 Oct 2019 20:22

Tgt Ion: 49 Resp: 49143
 Ion Ratio Lower Upper
 49 100
 129 2.1 0.0 4.2
 130 81.7 62.1 93.1



Abundance Ion 48.90 (48.60 to 49.60): VX0
 Ion 128.80 (128.50 to 129.50): V
 Ion 129.80 (129.50 to 130.50): V





#29

Tetrahydrofuran

Concen: 290.311 ug/l

RT: 5.12 min Scan# 694

Delta R.T. 0.00 min

Lab File: VX013065.D

Acq: 14 Oct 2019 20:22

Instrument :

MSVOA_X

ClientSampleId :


Tot Ion: 42 Resp: 233967

Ion Ratio Lower Upper

42 100

72 47.8 36.9 55.3




71 43.6 33.5 50.3



Abundance Ion 42.00 (41.70 to 42.70): VX0

Ion 72.00 (71.70 to 72.70): VX0

Ion 71.00 (70.70 to 71.70): VX0



#30

Chloroform

Concen: 53.877 ug/l

RT: 5.20 min Scan# 708

Delta R.T. 0.01 min

Lab File: VX013065.D



Acq: 14 Oct 2019 20:22

Tgt Ion: 83 Resp: 153676

Ion Ratio Lower Upper


83 100


85 65.3 51.5 77.3



Abundance Ion 82.90 (82.60 to 83.60): VX0

Ion 84.90 (84.60 to 85.60): VX0





#31

Cyclohexane

Concen: 44.155 ug/l

RT: 5.57 min Scan# 768

Delta R.T. 0.00 min

Lab File: VX013065.D

Acq: 14 Oct 2019 20:22

Instrument :
MSVOA_X
ClientSampleId :





Tot Ion: 56 Resp: 114773

Ion Ratio Lower Upper

56 100

69 31.9 25.2 37.8

84 85.4 71.3 106.9



#32

1,1,1-Trichloroethane

Concen: 51.517 ug/l

RT: 5.48 min Scan# 754

Delta R.T. -0.01 min

Lab File: VX013065.D

Acq: 14 Oct 2019 20:22




Tgt Ion: 97 Resp: 124505


Ion Ratio Lower Upper

97 100

99 65.0 52.3 78.5

61 42.7 35.0 52.4









#33
 1,2-Dichloroethane-d4
 Concen: 57.141 ug/l
 RT: 6.05 min Scan# 847
 Delta R.T. -0.01 min
 Lab File: VX013065.D
 Acq: 14 Oct 2019 20:22




Instrument :
 MSVOA_X
 ClientSampleId :


Tot Ion: 65 Resp: 102335
 Ion Ratio Lower Upper
 65 100
 67 52.8 0.0 104.4



#34
 1,4-Difluorobenzene
 Concen: 50.000 ug/l
 RT: 6.85 min Scan# 978
 Delta R.T. 0.00 min
 Lab File: VX013065.D
 Acq: 14 Oct 2019 20:22

Tgt Ion: 114 Resp: 251722
 Ion Ratio Lower Upper
 114 100
 63 20.6 0.0 40.8
 88 16.8 0.0 33.8





#35

Dibromofluoromethane

Concen: 55.960 ug/l

RT: 5.49 min Scan# 755

Delta R.T. 0.00 min

Lab File: VX013065.D

Acq: 14 Oct 2019 20:22

Instrument :

MSVOA_X

ClientSampleId :


Tot Ion:113 Resp: 80848

Ion Ratio Lower Upper

113 100

111 103.9 83.2 124.8




192 19.6 15.4 23.2



Abundance Ion 112.80 (112.50 to 113.50): V

Ion 110.80 (110.50 to 111.50): V

Ion 191.70 (191.40 to 192.40): V



#36

1,1-Dichloropropene

Concen: 46.548 ug/l

RT: 5.79 min Scan# 804

Delta R.T. 0.00 min

Lab File: VX013065.D

Acq: 14 Oct 2019 20:22


Tgt Ion: 75 Resp: 103018

Ion Ratio Lower Upper

75 100

110 34.9 17.6 52.9



77 31.1 24.4 36.6




Abundance Ion 74.90 (74.60 to 75.60): VX0

Ion 109.90 (109.60 to 110.60): V

Ion 76.90 (76.60 to 77.60): VX0









#37
Ethyl Acetate
Concen: 56.032 ug/l
RT: 4.82 min Scan# 645
Delta R.T. 0.00 min
Lab File: VX013065.D
Acq: 14 Oct 2019 20:22




Instrument :
MSVOA_X
ClientSampleId :


Tot Ion: 43 Resp: 138664
Ion Ratio Lower Upper
43 100
61 14.3 11.4 17.0
70 12.6 9.8 14.6



#38
Carbon Tetrachloride
Concen: 47.746 ug/l
RT: 5.77 min Scan# 801
Delta R.T. -0.01 min
Lab File: VX013065.D
Acq: 14 Oct 2019 20:22

Tgt Ion: 117 Resp: 102852
Ion Ratio Lower Upper
117 100
119 91.1 74.5 111.7
121 29.4 24.2 36.4









#39
Methvlcvclohexane
Concen: 43.292 ug/l
RT: 7.46 min Scan# 1078
Delta R.T. 0.00 min
Lab File: VX013065.D
Acq: 14 Oct 2019 20:22

Instrument :
MSVOA_X
ClientSampleId :

Tot Ion: 83 Resp: 116194
Ion Ratio Lower Upper
83 100
55 72.7 60.2 90.4
98 49.2 38.5 57.7




Abundance Ion 83.00 (82.70 to 83.70): VX0
Ion 55.00 (54.70 to 55.70): VX0
Ion 98.00 (97.70 to 98.70): VX0






#40
Benzene
Concen: 52.229 ug/l
RT: 6.14 min Scan# 861
Delta R.T. 0.00 min
Lab File: VX013065.D
Acq: 14 Oct 2019 20:22

Tgt Ion: 78 Resp: 341396
Ion Ratio Lower Upper
78 100
77 23.1 19.0 28.4



Abundance Ion 78.00 (77.70 to 78.70): VX0
Ion 77.00 (76.70 to 77.70): VX0





#41

Methacrylonitrile

Concen: 54.967 ug/l

RT: 5.03 min Scan# 679

Delta R.T. -0.01 min

Lab File: VX013065.D

Acq: 14 Oct 2019 20:22

Instrument :

MSVOA_X

ClientSampled :

Tot Ion: 41 Resp: 74746


Ion Ratio Lower Upper

41 100

39 55.9 44.0 66.0

67 75.4 58.5 87.7

52 31.2 23.8 35.8






Abundance Ion 41.00 (40.70 to 41.70): VX0

Ion 39.00 (38.70 to 39.70): VX0

Ion 67.00 (66.70 to 67.70): VX0

Ion 52.00 (51.70 to 52.70): VX0



#42

1,2-Dichloroethane

Concen: 53.723 ug/l

RT: 6.19 min Scan# 869

Delta R.T. 0.00 min

Lab File: VX013065.D


Acq: 14 Oct 2019 20:22

Tgt Ion: 62 Resp: 126007

Ion Ratio Lower Upper



62 100


98 9.5 0.0 20.4



Abundance Ion 61.90 (61.60 to 62.60): VX0

Ion 97.90 (97.60 to 98.60): VX0





#43

Isopropyl Acetate

Concen: 54.974 ug/l

RT: 6.44 min Scan# 910

Delta R.T. 0.00 min


Lab File: VX013065.D

Acq: 14 Oct 2019 20:22

Instrument :
MSVOA_X
ClientSampleId :

Tot Ion: 43 Resp: 218357

Ion	Ratio	Lower	Upper
43	100		
61	20.8	16.6	24.8
87	14.0	11.0	16.6






Abundance

Ion 43.00 (42.70 to 43.70): VX0

Ion 61.00 (60.70 to 61.70): VX0

Ion 87.00 (86.70 to 87.70): VX0



#44

Trichloroethene

Concen: 46.778 ug/l

RT: 7.21 min Scan# 1037


Delta R.T. 0.00 min

Lab File: VX013065.D

Acq: 14 Oct 2019 20:22

Tgt Ion: 130 Resp: 84790



Ion	Ratio	Lower	Upper
130	100		
95	100.5	0.0	189.2




Abundance

Ion 129.80 (129.50 to 130.50): V

Ion 94.90 (94.60 to 95.60): VX0









#45
 1,2-Dichloropropane
 Concen: 53.708 ug/l
 RT: 7.51 min Scan# 1086
 Delta R.T. 0.00 min
 Lab File: VX013065.D
 Acq: 14 Oct 2019 20:22




Instrument :
 MSVOA_X
 ClientSampled :


Tot Ion: 63 Resp: 89475
 Ion Ratio Lower Upper
 63 100
 65 31.9 24.6 37.0



#46
 Dibromomethane
 Concen: 54.465 ug/l
 RT: 7.65 min Scan# 1110
 Delta R.T. 0.00 min
 Lab File: VX013065.D
 Acq: 14 Oct 2019 20:22

Tgt Ion: 93 Resp: 60454
 Ion Ratio Lower Upper
 93 100
 95 85.2 66.3 99.5
 174 101.2 77.8 116.6





#47

Bromodichloromethane

Concen: 53.713 ug/l

RT: 7.89 min Scan# 1149

Delta R.T. 0.00 min

Lab File: VX013065.D

Acq: 14 Oct 2019 20:22

Instrument :

MSVOA_X

ClientSampleId :


Tot Ion: 83 Resp: 116348

Ion Ratio Lower Upper

83 100

85 64.7 52.4 78.6




127 9.3 6.6 10.0



Abundance Ion 82.90 (82.60 to 83.60): VX0

Ion 84.90 (84.60 to 85.60): VX0

Ion 126.80 (126.50 to 127.50): V



#48

Methyl methacrylate

Concen: 53.841 ug/l

RT: 7.76 min Scan# 1128

Delta R.T. 0.00 min

Lab File: VX013065.D

Acq: 14 Oct 2019 20:22


Tgt Ion: 41 Resp: 104215

Ion Ratio Lower Upper

41 100

69 89.6 68.7 103.1



39 57.5 44.3 66.5




Abundance Ion 41.00 (40.70 to 41.70): VX0

Ion 69.00 (68.70 to 69.70): VX0

Ion 39.00 (38.70 to 39.70): VX0









#49
1,4-Dioxane
Concen: 1069.602 ug/l
RT: 7.73 min Scan# 1123
Delta R.T. 0.00 min
Lab File: VX013065.D
Acq: 14 Oct 2019 20:22

Instrument :
MSVOA_X
ClientSampled :

Tot Ion	88	Resp	55690
Ion	Ratio	Lower	Upper
88	100		
43	30.9	25.0	37.6
58	67.8	55.4	83.2




Abundance Ion 88.00 (87.70 to 88.70): VX0
Ion 43.00 (42.70 to 43.70): VX0
Ion 58.00 (57.70 to 58.70): VX0






#50
Toluene-d8
Concen: 52.493 ug/l
RT: 8.71 min Scan# 1283
Delta R.T. 0.00 min
Lab File: VX013065.D
Acq: 14 Oct 2019 20:22

Tgt Ion	98	Resp	295793
Ion	Ratio	Lower	Upper
98	100		
100	67.5	53.4	80.2



Abundance Ion 98.00 (97.70 to 98.70): VX0
Ion 100.00 (99.70 to 100.70): VX0





#51

4-Methyl-2-Pentanone

Concen: 283.184 ug/l

RT: 8.64 min Scan# 1271

Delta R.T. 0.00 min

Lab File: VX013065.D

Acq: 14 Oct 2019 20:22





Instrument :
MSVOA_X
ClientSampled :

Tot Ion: 43 Resp: 701295

Ion Ratio Lower Upper

43 100

58 41.2 32.4 48.6



#52

Toluene

Concen: 51.138 ug/l

RT: 8.78 min Scan# 1295

Delta R.T. 0.00 min

Lab File: VX013065.D




Acq: 14 Oct 2019 20:22


Tgt Ion: 92 Resp: 213848

Ion Ratio Lower Upper

92 100

91 166.7 135.9 203.9









#53
 t-1,3-Dichloropropene
 Concen: 47.109 ug/l
 RT: 9.04 min Scan# 1337
 Delta R.T. 0.00 min
 Lab File: VX013065.D
 Acq: 14 Oct 2019 20:22




Instrument :
 MSVOA_X
 ClientSampleId :


Tot Ion: 75 Resp: 126991
 Ion Ratio Lower Upper
 75 100
 77 31.0 25.8 38.6



#54
 cis-1,3-Dichloropropene
 Concen: 52.716 ug/l
 RT: 8.43 min Scan# 1237
 Delta R.T. 0.00 min
 Lab File: VX013065.D
 Acq: 14 Oct 2019 20:22

Tgt Ion: 75 Resp: 140442
 Ion Ratio Lower Upper
 75 100
 77 31.5 25.8 38.8
 39 46.8 38.2 57.2









#55
 1,1,2-Trichloroethane
 Concen: 55.872 ug/l
 RT: 9.21 min Scan# 1365
 Delta R.T. 0.00 min
 Lab File: VX013065.D
 Acq: 14 Oct 2019 20:22




Instrument :
 MSVOA_X
 ClientSampleId :


Tot Ion:	97	Resp:	92464
Ion	Ratio	Lower	Upper
97	100		
83	86.1	66.5	99.7
85	53.4	43.1	64.7
99	63.8	49.2	73.8



#56
 Ethyl methacrylate
 Concen: 51.034 ug/l
 RT: 9.17 min Scan# 1359
 Delta R.T. 0.00 min
 Lab File: VX013065.D
 Acq: 14 Oct 2019 20:22

Tgt Ion:	69	Resp:	151354
Ion	Ratio	Lower	Upper
69	100		
41	64.0	52.1	78.1
39	33.1	27.8	41.8









#57
 1,3-Dichloropropane
 Concen: 54.342 ug/l
 RT: 9.37 min Scan# 1391
 Delta R.T. 0.00 min
 Lab File: VX013065.D
 Acq: 14 Oct 2019 20:22




Instrument :
 MSVOA_X
 ClientSampleId :


Tot Ion: 76 Resp: 155567
 Ion Ratio Lower Upper
 76 100
 78 32.0 26.1 39.1



#58
 2-Chloroethyl Vinyl ether
 Concen: 282.175 ug/l
 RT: 8.31 min Scan# 1217
 Delta R.T. 0.00 min
 Lab File: VX013065.D
 Acq: 14 Oct 2019 20:22

Tgt Ion: 63 Resp: 343053
 Ion Ratio Lower Upper
 63 100
 106 30.6 23.8 35.8









#59
2-Hexanone
Concen: 283.278 ug/l
RT: 9.49 min Scan# 1411
Delta R.T. 0.00 min
Lab File: VX013065.D
Acq: 14 Oct 2019 20:22




Instrument :
MSVOA_X
ClientSampled :


Tot Ion: 43 Resp: 546095
Ion Ratio Lower Upper
43 100
58 56.3 28.1 84.3



#60
Dibromochloromethane
Concen: 49.838 ug/l
RT: 9.58 min Scan# 1426
Delta R.T. 0.00 min
Lab File: VX013065.D
Acq: 14 Oct 2019 20:22

Tgt Ion: 129 Resp: 91772
Ion Ratio Lower Upper
129 100
127 77.4 38.9 116.7





#61

1,2-Dibromoethane

Concen: 54.104 ug/l

RT: 9.67 min Scan# 1440

Delta R.T. 0.00 min

Lab File: VX013065.D

Acq: 14 Oct 2019 20:22

Instrument :

MSVOA_X





ClientSampleId :

Tot Ion:107 Resp: 94189

Ion Ratio Lower Upper

107 100

109 95.5 75.8 113.6



#62

4-Bromofluorobenzene

Concen: 50.996 ug/l

RT: 11.14 min Scan# 1681

Delta R.T. 0.00 min

Lab File: VX013065.D

Acq: 14 Oct 2019 20:22




Tgt Ion: 95 Resp: 111654


Ion Ratio Lower Upper

95 100

174 76.3 0.0 143.4

176 75.1 0.0 136.0









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.11 min Scan# 1513
Delta R.T. 0.00 min
Lab File: VX013065.D
Acq: 14 Oct 2019 20:22

Instrument :
MSVOA_X
ClientSampleId :

Tot Ion:117 Resp: 231619
Ion Ratio Lower Upper
117 100
82 56.1 44.1 66.1
119 31.8 25.8 38.8




Abundance Ion 116.90 (116.60 to 117.60): V
Ion 82.00 (81.70 to 82.70): VX0
Ion 118.90 (118.60 to 119.60): V






#64
Tetrachloroethene
Concen: 49.975 ug/l
RT: 9.33 min Scan# 1385
Delta R.T. 0.00 min
Lab File: VX013065.D
Acq: 14 Oct 2019 20:22

Tgt Ion:164 Resp: 80285
Ion Ratio Lower Upper
164 100
166 130.8 96.9 145.3
129 95.5 77.9 116.9
131 94.3 72.7 109.1



Abundance Ion 163.80 (163.50 to 164.50): V
Ion 165.80 (165.50 to 166.50): V
Ion 128.80 (128.50 to 129.50): V
Ion 130.80 (130.50 to 131.50): V





#65

Chlorobenzene

Concen: 49.511 ug/l

RT: 10.14 min Scan# 1517

Delta R.T. 0.00 min

Lab File: VX013065.D

Acq: 14 Oct 2019 20:22

Instrument :

MSVOA_X





ClientSampleId :

Tot Ion:112 Resp: 227730

Ion Ratio Lower Upper

112 100

114 32.4 25.7 38.5



#66

1,1,1,2-Tetrachloroethane

Concen: 53.355 ug/l

RT: 10.21 min Scan# 1530

Delta R.T. 0.00 min

Lab File: VX013065.D

Acq: 14 Oct 2019 20:22




Tgt Ion:131 Resp: 85089


Ion Ratio Lower Upper

131 100

133 96.4 47.2 141.6

119 65.8 32.6 98.0









#67
Ethyl Benzene
Concen: 48.609 ug/l
RT: 10.25 min Scan# 1535
Delta R.T. 0.00 min
Lab File: VX013065.D
Acq: 14 Oct 2019 20:22




Instrument :
MSVOA_X
ClientSampleId :


Tot Ion: 91 Resp: 397141
Ion Ratio Lower Upper
91 100
106 31.9 23.9 35.9



#68
m/p-Xylenes
Concen: 99.781 ug/l
RT: 10.36 min Scan# 1553
Delta R.T. 0.00 min
Lab File: VX013065.D
Acq: 14 Oct 2019 20:22

Tgt Ion: 106 Resp: 304554
Ion Ratio Lower Upper
106 100
91 204.7 165.3 247.9









#69
o-Xylene
Concen: 50.366 ug/l
RT: 10.70 min Scan# 1609
Delta R.T. 0.00 min
Lab File: VX013065.D
Acq: 14 Oct 2019 20:22

Instrument :
MSVOA_X
ClientSampleId :

Tot Ion:106 Resp: 149913
Ion Ratio Lower Upper
106 100
91 219.3 109.1 327.3




Abundance Ion 106.00 (105.70 to 106.70): V
Ion 91.00 (90.70 to 91.70): VX0






#70
Styrene
Concen: 52.055 ug/l
RT: 10.71 min Scan# 1611
Delta R.T. 0.00 min
Lab File: VX013065.D
Acq: 14 Oct 2019 20:22

Tgt Ion:104 Resp: 263399
Ion Ratio Lower Upper
104 100
78 52.1 41.2 61.8
103 54.4 43.0 64.4



Abundance Ion 104.00 (103.70 to 104.70): V
Ion 78.00 (77.70 to 78.70): VX0
Ion 103.00 (102.70 to 103.70): V





#71

Bromoform

Concen: 47.110 ug/l

RT: 10.86 min Scan# 1635

Delta R.T. 0.00 min

Lab File: VX013065.D

Acq: 14 Oct 2019 20:22

Instrument :

MSVOA_X

ClientSampleId :


Tot Ion:173 Resp: 63741

Ion Ratio Lower Upper




173 100

175 47.7 24.1 72.3

254 0.2 0.1 0.1#



Abundance Ion 172.70 (172.40 to 173.40): V
Ion 174.70 (174.40 to 175.40): V
Ion 254.40 (254.10 to 255.10): V



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 12.07 min Scan# 1835

Delta R.T. 0.00 min

Lab File: VX013065.D

Acq: 14 Oct 2019 20:22


Tgt Ion:152 Resp: 111222

Ion Ratio Lower Upper



152 100


115 84.1 44.5 133.5

150 169.3 0.0 353.4



Abundance Ion 151.90 (151.60 to 152.60): V
Ion 114.90 (114.60 to 115.60): V
Ion 149.90 (149.60 to 150.60): V





#73

Isopropylbenzene

Concen: 46.538 ug/l

RT: 11.01 min Scan# 1661

Delta R.T. 0.00 min

Lab File: VX013065.D

Acq: 14 Oct 2019 20:22

Instrument :

MSVOA_X


ClientSampleId :

Tot Ion:105 Resp: 380550

Ion Ratio Lower Upper

105 100

120 26.3 13.0 39.0



Abundance Ion 105.00 (104.70 to 105.70): V

Ion 120.00 (119.70 to 120.70): V

300000

200000



100000

0

11.01

11.10

Time-->



#74

N-ethyl acetate

Concen: 52.823 ug/l

RT: 10.89 min Scan# 1641

Delta R.T. 0.00 min

Lab File: VX013065.D

Acq: 14 Oct 2019 20:22

Tgt Ion: 43 Resp: 186030


Ion Ratio Lower Upper

43 100

70 45.0 36.4 54.6

55 25.2 20.2 30.2

61 25.3 20.5 30.7



Abundance Ion 43.00 (42.70 to 43.70): VX0

Ion 70.00 (69.70 to 70.70): VX0

Ion 55.00 (54.70 to 55.70): VX0

Ion 61.00 (60.70 to 61.70): VX0

250000

200000

150000

100000


50000


0

10.89

11.00

Time-->





#75

1,1,2,2-Tetrachloroethane

Concen: 52.147 ug/l

RT: 11.26 min Scan# 1702

Delta R.T. 0.00 min

Lab File: VX013065.D

Acq: 14 Oct 2019 20:22

Instrument :

MSVOA_X

ClientSampleId :


Tot Ion: 83 Resp: 141827

Ion Ratio Lower Upper

83 100

131 10.4 5.1 15.3




85 65.2 32.1 96.5



Abundance Ion 82.90 (82.60 to 83.60): VX0

Ion 130.80 (130.50 to 131.50): V

Ion 84.90 (84.60 to 85.60): VX0



#76

1,2,3-Trichloropropane

Concen: 64.492 ug/l

RT: 11.29 min Scan# 1707

Delta R.T. 0.01 min

Lab File: VX013065.D


Acq: 14 Oct 2019 20:22

Tgt Ion: 75 Resp: 159094

Ion Ratio Lower Upper



75 100


77 32.9 22.9 68.8



Abundance Ion 74.90 (74.60 to 75.60): VX0

Ion 77.00 (76.70 to 77.70): VX0





#77

Bromobenzene

Concen: 50.043 ug/l

RT: 11.25 min Scan# 1700

Delta R.T. 0.00 min

Lab File: VX013065.D

Acq: 14 Oct 2019 20:22

Instrument :

MSVOA_X

ClientSampleId :


Tot Ion:156 Resp: 96431

Ion Ratio Lower Upper

156 100

77 161.2 83.7 251.0




158 95.9 48.6 145.9



Abundance Ion 155.80 (155.50 to 156.50): V

Ion 77.00 (76.70 to 77.70): VX0

Ion 157.80 (157.50 to 158.50): V



#78

n-propylbenzene

Concen: 46.971 ug/l

RT: 11.35 min Scan# 1717

Delta R.T. 0.00 min

Lab File: VX013065.D


Acq: 14 Oct 2019 20:22

Tgt Ion: 91 Resp: 422645

Ion Ratio Lower Upper



91 100


120 22.9 11.7 35.1



Abundance Ion 91.00 (90.70 to 91.70): VX0

Ion 120.00 (119.70 to 120.70): V









#79
 2-Chlorotoluene
 Concen: 47.143 ug/l
 RT: 11.42 min Scan# 1727
 Delta R.T. 0.00 min
 Lab File: VX013065.D
 Acq: 14 Oct 2019 20:22




Instrument :
 MSVOA_X
 ClientSampleId :


Tot Ion: 91 Resp: 266592
 Ion Ratio Lower Upper
 91 100
 126 32.2 16.3 48.8



#80
 1,3,5-Trimethylbenzene
 Concen: 48.313 ug/l
 RT: 11.50 min Scan# 1741
 Delta R.T. 0.00 min
 Lab File: VX013065.D
 Acq: 14 Oct 2019 20:22

Tgt Ion: 105 Resp: 329880
 Ion Ratio Lower Upper
 105 100
 120 48.4 24.8 74.3









#81
trans-1,4-Dichloro-2-butene
Concen: 46.088 ug/l
RT: 11.07 min Scan# 1670
Delta R.T. 0.00 min
Lab File: VX013065.D
Acq: 14 Oct 2019 20:22




Instrument :
MSVOA_X
ClientSampled :


Tot Ion: 75 Resp: 43880
Ion Ratio Lower Upper
75 100
53 97.5 78.5 117.7
89 46.7 37.4 56.2



#82
4-Chlorotoluene
Concen: 48.557 ug/l
RT: 11.51 min Scan# 1742
Delta R.T. 0.00 min
Lab File: VX013065.D
Acq: 14 Oct 2019 20:22

Tgt Ion: 91 Resp: 314731
Ion Ratio Lower Upper
91 100
126 28.1 14.2 42.6









#83
tert-Butylbenzene
Concen: 47.422 ug/l
RT: 11.76 min Scan# 1784
Delta R.T. 0.00 min
Lab File: VX013065.D
Acq: 14 Oct 2019 20:22




Instrument :
MSVOA_X
ClientSampled :


Tot Ion:119 Resp: 315865
Ion Ratio Lower Upper
119 100
91 58.2 29.3 87.9
134 22.1 11.5 34.4



#84
1,2,4-Trimethylbenzene
Concen: 49.606 ug/l
RT: 11.80 min Scan# 1790
Delta R.T. -0.01 min
Lab File: VX013065.D
Acq: 14 Oct 2019 20:22

Tgt Ion:105 Resp: 343039
Ion Ratio Lower Upper
105 100
120 42.5 21.8 65.3





#85

sec-Butylbenzene

Concen: 45.504 ug/l

RT: 11.94 min Scan# 1813

Delta R.T. 0.00 min

Lab File: VX013065.D

Acq: 14 Oct 2019 20:22

Instrument :

MSVOA_X


ClientSampleId :

Tot Ion:105 Resp: 351141

Ion Ratio Lower Upper

105 100

134 19.3 10.0 30.0



Abundance Ion 105.00 (104.70 to 105.70): V

Ion 134.00 (133.70 to 134.70): V

11.94



300000

200000

100000

0

Time-->



#86

p-Isopropyltoluene

Concen: 45.604 ug/l

RT: 12.06 min Scan# 1833

Delta R.T. 0.00 min

Lab File: VX013065.D

Acq: 14 Oct 2019 20:22


Tgt Ion:119 Resp: 326378

Ion Ratio Lower Upper

119 100

134 25.3 12.6 37.8

91 25.6 12.4 37.4



Abundance Ion 119.00 (118.70 to 119.70): V

Ion 134.00 (133.70 to 134.70): V

Ion 91.00 (90.70 to 91.70): VX0

12.06


300000


200000

100000

0

Time-->









#87
 1,3-Dichlorobenzene
 Concen: 48.198 ug/l
 RT: 12.02 min Scan# 1826
 Delta R.T. 0.00 min
 Lab File: VX013065.D
 Acq: 14 Oct 2019 20:22

Instrument :
 MSVOA_X
 ClientSampleId :

Tot Ion:146 Resp: 172471
 Ion Ratio Lower Upper
 146 100
 111 41.7 21.1 63.4
 148 63.4 31.6 95.0




Abundance Ion 145.90 (145.60 to 146.60): V
 Ion 110.90 (110.60 to 111.60): V
 Ion 147.90 (147.60 to 148.60): V






#88
 1,4-Dichlorobenzene
 Concen: 48.115 ug/l
 RT: 12.09 min Scan# 1838
 Delta R.T. 0.00 min
 Lab File: VX013065.D
 Acq: 14 Oct 2019 20:22

Tgt Ion:146 Resp: 174370
 Ion Ratio Lower Upper
 146 100
 111 42.5 21.4 64.3
 148 62.2 32.1 96.5



Abundance Ion 145.90 (145.60 to 146.60): V
 Ion 110.90 (110.60 to 111.60): V
 Ion 147.90 (147.60 to 148.60): V









#89
n-Butylbenzene
Concen: 43.383 ug/l
RT: 12.39 min Scan# 1886
Delta R.T. 0.00 min
Lab File: VX013065.D
Acq: 14 Oct 2019 20:22




Instrument :
MSVOA_X
ClientSampled :


Tot Ion: 91 Resp: 260945
Ion Ratio Lower Upper
91 100
92 56.8 27.5 82.5
134 27.3 12.9 38.7



#90
Hexachloroethane
Concen: 40.762 ug/l
RT: 12.59 min Scan# 1920
Delta R.T. 0.00 min
Lab File: VX013065.D
Acq: 14 Oct 2019 20:22

Tgt Ion: 117 Resp: 53490
Ion Ratio Lower Upper
117 100
201 77.6 36.9 110.7





#91

1,2-Dichlorobenzene

Concen: 49.972 ug/l

RT: 12.39 min Scan# 1886

Delta R.T. 0.00 min

Lab File: VX013065.D

Acq: 14 Oct 2019 20:22

Instrument :

MSVOA_X

ClientSampleId :


Tot Ion:146 Resp: 175307

Ion Ratio Lower Upper

146 100

111 44.6 22.7 68.0




148 63.8 31.6 94.7



Abundance Ion 145.90 (145.60 to 146.60): V

Ion 110.90 (110.60 to 111.60): V

Ion 147.90 (147.60 to 148.60): V



#92

1,2-Dibromo-3-Chloropropane

Concen: 55.033 ug/l

RT: 12.99 min Scan# 1986

Delta R.T. 0.00 min

Lab File: VX013065.D

Acq: 14 Oct 2019 20:22


Tgt Ion: 75 Resp: 35082

Ion Ratio Lower Upper

75 100

155 84.4 43.0 128.8



157 109.0 54.1 162.3




Abundance Ion 74.90 (74.60 to 75.60): VX0

Ion 154.80 (154.50 to 155.50): V

Ion 156.80 (156.50 to 157.50): V









#93
 1,2,4-Trichlorobenzene
 Concen: 48.152 ug/l
 RT: 13.64 min Scan# 2092
 Delta R.T. 0.00 min
 Lab File: VX013065.D
 Acq: 14 Oct 2019 20:22

Instrument :
 MSVOA_X
 ClientSampleId :

Tot Ion:180 Resp: 105816
 Ion Ratio Lower Upper
 180 100
 182 94.2 46.8 140.4
 145 32.2 16.2 48.6




Abundance Ion 179.80 (179.50 to 180.50): V
 Ion 181.80 (181.50 to 182.50): V
 Ion 144.80 (144.50 to 145.50): V






#94
 Hexachlorobutadiene
 Concen: 41.337 ug/l
 RT: 13.78 min Scan# 2114
 Delta R.T. -0.01 min
 Lab File: VX013065.D
 Acq: 14 Oct 2019 20:22

Tgt Ion:225 Resp: 41923
 Ion Ratio Lower Upper
 225 100
 223 61.9 31.1 93.2
 227 62.4 32.3 96.9



Abundance Ion 224.70 (224.40 to 225.40): V
 Ion 222.70 (222.40 to 223.40): V
 Ion 226.70 (226.40 to 227.40): V





#95

Naphthalene

Concen: 52.289 ug/l

RT: 13.83 min Scan# 2123

Delta R.T. 0.00 min

Lab File: VX013065.D

Acq: 14 Oct 2019 20:22

Instrument :

MSVOA_X

ClientSampleId :


Tot Ion:128 Resp: 398331

Ion Ratio Lower Upper

128 100

127 12.9 10.2 15.4




129 10.8 8.8 13.2



Abundance Ion 128.00 (127.70 to 128.70): V

Ion 127.00 (126.70 to 127.70): V

Ion 129.00 (128.70 to 129.70): V



#96

1,2,3-Trichlorobenzene

Concen: 47.100 ug/l

RT: 14.01 min Scan# 2153

Delta R.T. 0.00 min

Lab File: VX013065.D

Acq: 14 Oct 2019 20:22


Tgt Ion:180 Resp: 106960

Ion Ratio Lower Upper

180 100

182 95.0 47.5 142.5

145 34.8 17.4 52.3



Abundance Ion 179.80 (179.50 to 180.50): V

Ion 181.80 (181.50 to 182.50): V

Ion 144.90 (144.60 to 145.60): V

