

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX101419\
 Data File : VX013044.D
 Acq On : 14 Oct 2019 12:13
 Operator : JC/SP
 Sample : K5313-12
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FA19-JBW7338A

Quant Time: Oct 15 03:22:35 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X100819W.M
 Quant Title : SW846 8260
 QLast Update : Fri Oct 11 09:48:22 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	184210	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	302835	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	261447	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	108095	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	113288	49.20	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.40%	
35) Dibromofluoromethane	5.49	113	88724	51.05	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.10%	
50) Toluene-d8	8.71	98	356342	52.57	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.14%	
62) 4-Bromofluorobenzene	11.14	95	122591	46.54	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.08%	
Target Compounds						
					Qvalue	
4) Vinyl Chloride	1.40	62	6234	2.956	ug/l	92
16) Acetone	2.43	43	11093	9.141	ug/l	94
27) cis-1,2-Dichloroethene	4.59	96	5580	2.426	ug/l	90
31) Cyclohexane	5.57	56	10671	3.193	ug/l	96
39) Methylcyclohexane	7.45	83	2431	0.753	ug/l #	85
40) Benzene	6.14	78	53706	6.829	ug/l	99
44) Trichloroethene	7.21	130	3345	1.534	ug/l	87
73) Isopropylbenzene	11.01	105	12262	1.543	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX101419\
Data File : VX013044.D
Acq On : 14 Oct 2019 12:13
Operator : JC/SP
Sample : K5313-12
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FA19-JBW7338A

Quant Time: Oct 15 03:22:35 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X100819W.M
Quant Title : SW846 8260
QLast Update : Fri Oct 11 09:48:22 2019
Response via : Initial Calibration

