

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX101419\
 Data File : VX013049.D
 Acq On : 14 Oct 2019 14:09
 Operator : JC/SP
 Sample : K5312-05
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 S49-0549

Quant Time: Oct 15 09:15:04 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X100819W.M
 Quant Title : SW846 8260
 QLast Update : Fri Oct 11 09:48:22 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	180602	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	296474	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	247380	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	98053	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	109706	48.60	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.20%	
35) Dibromofluoromethane	5.49	113	88414	51.96	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.92%	
50) Toluene-d8	8.71	98	342138	51.55	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.10%	
62) 4-Bromofluorobenzene	11.14	95	111715	43.32	ug/l	0.00
Spiked Amount 50.000			Recovery	=	86.64%	
Target Compounds						
16) Acetone	2.43	43	2227	1.872	ug/l #	88
27) cis-1,2-Dichloroethene	4.61	96	3203	1.421	ug/l	86
44) Trichloroethene	7.22	130	1332	0.624	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX101419\
Data File : VX013049.D
Acq On : 14 Oct 2019 14:09
Operator : JC/SP
Sample : K5312-05
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
S49-0549

Quant Time: Oct 15 09:15:04 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X100819W.M
Quant Title : SW846 8260
QLast Update : Fri Oct 11 09:48:22 2019
Response via : Initial Calibration

