

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX101419\
 Data File : VX013053.D
 Acq On : 14 Oct 2019 15:43
 Operator : JC/SP
 Sample : K5336-01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FA19-USMW025

Quant Time: Oct 15 05:04:01 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X100819W.M
 Quant Title : SW846 8260
 QLast Update : Fri Oct 11 09:48:22 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	176359	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	287498	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	239115	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	88876	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	108137	49.06	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.12%	
35) Dibromofluoromethane	5.49	113	86682	52.53	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.06%	
50) Toluene-d8	8.71	98	332232	51.62	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.24%	
62) 4-Bromofluorobenzene	11.14	95	103252	41.29	ug/l	0.00
Spiked Amount 50.000			Recovery	=	82.58%	
Target Compounds						
					Qvalue	
4) Vinyl Chloride	1.40	62	2224	1.101	ug/l	96
16) Acetone	2.44	43	8524	7.336	ug/l	99
27) cis-1,2-Dichloroethene	4.58	96	2675	1.215	ug/l #	72
37) Ethyl Acetate	4.82	43	23017	8.143	ug/l	98
44) Trichloroethene	7.21	130	6423	3.103	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX101419\
Data File : VX013053.D
Acq On : 14 Oct 2019 15:43
Operator : JC/SP
Sample : K5336-01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FA19-USMW025

Quant Time: Oct 15 05:04:01 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X100819W.M
Quant Title : SW846 8260
QLast Update : Fri Oct 11 09:48:22 2019
Response via : Initial Calibration

