

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX101419\
 Data File : VX013056.D
 Acq On : 14 Oct 2019 16:53
 Operator : JC/SP
 Sample : K5336-04
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FA19-JMW6001

Quant Time: Oct 15 12:19:55 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X100819W.M
 Quant Title : SW846 8260
 QLast Update : Fri Oct 11 09:48:22 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	172244	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	278664	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	229334	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	80433	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	106334	49.39	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.78%	
35) Dibromofluoromethane	5.49	113	84278	52.69	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.38%	
50) Toluene-d8	8.71	98	321181	51.49	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.98%	
62) 4-Bromofluorobenzene	11.14	95	99351	40.99	ug/l	0.00
Spiked Amount 50.000			Recovery	=	81.98%	
Target Compounds						
					Qvalue	
4) Vinyl Chloride	1.40	62	2834	1.437	ug/l	99
16) Acetone	2.44	43	5443	4.797	ug/l	94
21) trans-1,2-Dichloroethene	3.16	96	1107	0.593	ug/l #	83
27) cis-1,2-Dichloroethene	4.59	96	175378	81.554	ug/l	90
37) Ethyl Acetate	4.82	43	209310	76.402	ug/l	98
44) Trichloroethene	7.21	130	82069	40.899	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX101419\
Data File : VX013056.D
Acq On : 14 Oct 2019 16:53
Operator : JC/SP
Sample : K5336-04
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FA19-JMW6001

Quant Time: Oct 15 12:19:55 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X100819W.M
Quant Title : SW846 8260
QLast Update : Fri Oct 11 09:48:22 2019
Response via : Initial Calibration

