

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX101419\
 Data File : VX013063.D
 Acq On : 14 Oct 2019 19:36
 Operator : JC/SP
 Sample : K5266-07ME
 Misc : 5.00g/10mL/100uL/5.0mL/MSVOA_X/MEOH
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 20190856-AMENDED-COMP-CME

Quant Time: Oct 15 12:29:24 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X100819W.M
 Quant Title : SW846 8260
 QLast Update : Fri Oct 11 09:48:22 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	166999	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	265793	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	229243	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	99957	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	99123	47.49	ug/l	0.00
Spiked Amount	50.000		Recovery	=	94.98%	
35) Dibromofluoromethane	5.49	113	74739	48.99	ug/l	0.00
Spiked Amount	50.000		Recovery	=	97.98%	
50) Toluene-d8	8.71	98	310689	52.22	ug/l	0.00
Spiked Amount	50.000		Recovery	=	104.44%	
62) 4-Bromofluorobenzene	11.14	95	105905	45.81	ug/l	0.00
Spiked Amount	50.000		Recovery	=	91.62%	
Target Compounds						
4) Vinyl Chloride	1.41	62	18039	9.434	ug/l	93
12) 1,1-Dichloroethene	2.33	96	6969	4.130	ug/l	92
24) 1,1-Dichloroethane	3.68	63	4456	1.375	ug/l #	94
27) cis-1,2-Dichloroethene	4.58	96	65101	31.224	ug/l	87
39) Methylcyclohexane	7.45	83	6500	2.294	ug/l	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX101419\
Data File : VX013063.D
Acq On : 14 Oct 2019 19:36
Operator : JC/SP
Sample : K5266-07ME
Misc : 5.00µ/10mL/100µL/5.0mL/MSVOA_X/MEOH
ALS Vial : 28 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
20190856-AMENDED-COMP-CME

Quant Time: Oct 15 12:29:24 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X100819W.M
Quant Title : SW846 8260
QLast Update : Fri Oct 11 09:48:22 2019
Response via : Initial Calibration

