

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101421\
 Data File : VX024817.D
 Acq On : 14 Oct 2021 16:37
 Operator : JC/MD
 Sample : VX1014WBSD01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX1014WBSD01

Quant Time: Oct 15 00:02:54 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X101321W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 14 03:51:38 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	158325	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	247033	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	237385	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	118203	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	98464	50.599	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	101.200%
35) Dibromofluoromethane	5.391	113	74196	50.452	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	100.900%
50) Toluene-d8	8.653	98	275365	50.619	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	101.240%
62) 4-Bromofluorobenzene	11.085	95	105610	49.361	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	98.720%
Target Compounds						
						Qvalue
2) Dichlorodifluoromethane	1.166	85	29809	17.675	ug/l	98
3) Chloromethane	1.294	50	25862	17.610	ug/l	98
4) Vinyl Chloride	1.374	62	26519	17.428	ug/l	96
5) Bromomethane	1.605	94	18840	21.122	ug/l	99
6) Chloroethane	1.684	64	16258	17.364	ug/l	96
7) Trichlorofluoromethane	1.886	101	44869	17.259	ug/l	96
8) Diethyl Ether	2.136	74	15840	18.556	ug/l	95
9) 1,1,2-Trichlorotrifluo...	2.331	101	26613	17.408	ug/l	99
10) Methyl Iodide	2.453	142	30711	16.544	ug/l	100
11) Tert butyl alcohol	2.977	59	39115	95.363	ug/l	99
12) 1,1-Dichloroethene	2.319	96	23802	17.231	ug/l	100
13) Acrolein	2.239	56	17560	103.099	ug/l	99
14) Allyl chloride	2.666	41	44599	17.877	ug/l	99
15) Acrylonitrile	3.068	53	88667	96.950	ug/l	100
16) Acetone	2.386	43	86150	73.732	ug/l	98
17) Carbon Disulfide	2.514	76	57034	15.804	ug/l	99
18) Methyl Acetate	2.709	43	63647	19.814	ug/l	100
19) Methyl tert-butyl Ether	3.123	73	93967	18.420	ug/l	99
20) Methylene Chloride	2.794	84	29983	18.168	ug/l	97
21) trans-1,2-Dichloroethene	3.093	96	26546	17.222	ug/l	98
22) Diisopropyl ether	3.763	45	93023	18.472	ug/l #	87
23) Vinyl Acetate	3.727	43	359384	92.925	ug/l	100
24) 1,1-Dichloroethane	3.611	63	51303	17.708	ug/l	98
25) 2-Butanone	4.568	43	129264	93.337	ug/l	98
26) 2,2-Dichloropropane	4.483	77	41556	16.495	ug/l	100
27) cis-1,2-Dichloroethene	4.495	96	31240	17.678	ug/l	99
28) Bromochloromethane	4.903	49	22323	18.449	ug/l	98
29) Tetrahydrofuran	5.019	42	83111	94.438	ug/l	99
30) Chloroform	5.098	83	54654	17.550	ug/l	98
31) Cyclohexane	5.476	56	44553	17.186	ug/l	97
32) 1,1,1-Trichloroethane	5.385	97	48361	17.395	ug/l	99
36) 1,1-Dichloropropene	5.696	75	37094	16.967	ug/l	98
37) Ethyl Acetate	4.727	43	47684	18.932	ug/l	99
38) Carbon Tetrachloride	5.684	117	42241	17.245	ug/l	97
39) Methylcyclohexane	7.385	83	47170	17.404	ug/l	98
40) Benzene	6.043	78	113007	17.985	ug/l	98

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101421\
 Data File : VX024817.D
 Acq On : 14 Oct 2021 16:37
 Operator : JC/MD
 Sample : VX1014WBSD01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX1014WBSD01

Quant Time: Oct 15 00:02:54 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X101321W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 14 03:51:38 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	4.928	41	26575	19.338	ug/l	98
42) 1,2-Dichloroethane	6.092	62	47608	18.402	ug/l	100
43) Isopropyl Acetate	6.348	43	73143	18.775	ug/l	99
44) Trichloroethene	7.129	130	31352	17.800	ug/l	99
45) 1,2-Dichloropropane	7.433	63	29495	18.048	ug/l	95
46) Dibromomethane	7.586	93	21967	18.572	ug/l	99
47) Bromodichloromethane	7.824	83	40473	18.098	ug/l	96
48) Methyl methacrylate	7.696	41	36122	19.012	ug/l	99
49) 1,4-Dioxane	7.665	88	15901	388.334	ug/l	98
51) 4-Methyl-2-Pentanone	8.580	43	249363	96.433	ug/l	100
52) Toluene	8.720	92	73206	18.005	ug/l	99
53) t-1,3-Dichloropropene	8.982	75	43785	17.620	ug/l	99
54) cis-1,3-Dichloropropene	8.372	75	46932	18.027	ug/l	98
55) 1,1,2-Trichloroethane	9.153	97	32310	18.807	ug/l	97
56) Ethyl methacrylate	9.122	69	47077	17.007	ug/l	98
57) 1,3-Dichloropropane	9.311	76	53741	19.188	ug/l	99
58) 2-Chloroethyl Vinyl ether	8.244	63	127299	106.019	ug/l	100
59) 2-Hexanone	9.433	43	197065	86.540	ug/l	99
60) Dibromochloromethane	9.525	129	30813	16.673	ug/l	99
61) 1,2-Dibromoethane	9.610	107	33857	18.335	ug/l	97
64) Tetrachloroethene	9.281	164	30580	17.602	ug/l	92
65) Chlorobenzene	10.079	112	80555	17.332	ug/l	99
66) 1,1,1,2-Tetrachloroethane	10.165	131	30777	17.848	ug/l	97
67) Ethyl Benzene	10.195	91	139208	17.185	ug/l	99
68) m/p-Xylenes	10.305	106	109732	35.112	ug/l	98
69) o-Xylene	10.646	106	54638	17.816	ug/l	96
70) Styrene	10.658	104	88802	16.059	ug/l	99
71) Bromoform	10.805	173	22144	16.431	ug/l #	97
73) Isopropylbenzene	10.963	105	142391	18.213	ug/l	99
74) N-amyl acetate	10.847	43	58825	17.961	ug/l	99
75) 1,1,2,2-Tetrachloroethane	11.213	83	49945	18.836	ug/l	99
76) 1,2,3-Trichloropropane	11.244	75	56775	23.023	ug/l	88
77) Bromobenzene	11.201	156	35394	17.619	ug/l	97
78) n-propylbenzene	11.305	91	163732	18.178	ug/l	99
79) 2-Chlorotoluene	11.366	91	101310	18.264	ug/l	100
80) 1,3,5-Trimethylbenzene	11.457	105	124607	18.789	ug/l	99
81) trans-1,4-Dichloro-2-b...	11.018	75	14149	16.177	ug/l	95
82) 4-Chlorotoluene	11.457	91	117519	18.230	ug/l	100
83) tert-Butylbenzene	11.719	119	117841	18.335	ug/l	99
84) 1,2,4-Trimethylbenzene	11.756	105	122779	18.365	ug/l	99
85) sec-Butylbenzene	11.896	105	151566	18.422	ug/l	99
86) p-Isopropyltoluene	12.012	119	128394	18.595	ug/l	99
87) 1,3-Dichlorobenzene	11.969	146	68521	18.093	ug/l	97
88) 1,4-Dichlorobenzene	12.042	146	68679	17.744	ug/l	99
89) n-Butylbenzene	12.335	91	108864	17.894	ug/l	99
90) Hexachloroethane	12.542	117	18547	16.741	ug/l	98
91) 1,2-Dichlorobenzene	12.341	146	68064	18.182	ug/l	98
92) 1,2-Dibromo-3-Chloropr...	12.945	75	11221	18.497	ug/l	99
93) 1,2,4-Trichlorobenzene	13.591	180	41124	18.000	ug/l	98
94) Hexachlorobutadiene	13.731	225	19202	17.408	ug/l	100
95) Naphthalene	13.780	128	136006	16.495	ug/l	100
96) 1,2,3-Trichlorobenzene	13.963	180	41522	18.085	ug/l	98


Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101421\
Data File : VX024817.D
Acq On : 14 Oct 2021 16:37
Operator : JC/MD
Sample : VX1014WBSD01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX1014WBSD01

Quant Time: Oct 15 00:02:54 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X101321W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 14 03:51:38 2021
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------





(#) = qualifier out of range (m) = manual integration (+) = signals summed



#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.556 min Scan# 733
Delta R.T. -0.006 min
Lab File: VX024817.D
Acq: 14 Oct 2021 16:37




Instrument :
MSVOA_X
ClientSampleId :
VX1014WBSD01


Tgt Ion:168 Resp: 158325
Ion Ratio Lower Upper
168 100
99 61.0 46.4 69.6



#2
Dichlorodifluoromethane
Concen: 17.675 ug/l
RT: 1.166 min Scan# 13
Delta R.T. -0.000 min
Lab File: VX024817.D
Acq: 14 Oct 2021 16:37

Tgt Ion: 85 Resp: 29809
Ion Ratio Lower Upper
85 100
87 33.2 16.0 47.9









#3
Chloromethane
Concen: 17.610 ug/l
RT: 1.294 min Scan# 34
Delta R.T. -0.000 min
Lab File: VX024817.D
Acq: 14 Oct 2021 16:37




Instrument :
MSVOA_X
ClientSampleId :
VX1014WBSD01


Tgt Ion: 50 Resp: 25862
Ion Ratio Lower Upper
50 100
52 33.4 25.9 38.9



#4
Vinyl Chloride
Concen: 17.428 ug/l
RT: 1.374 min Scan# 47
Delta R.T. -0.000 min
Lab File: VX024817.D
Acq: 14 Oct 2021 16:37

Tgt Ion: 62 Resp: 26519
Ion Ratio Lower Upper
62 100
64 30.4 26.0 39.0









#5
Bromomethane
Concen: 21.122 ug/l
RT: 1.605 min Scan# 85
Delta R.T. 0.006 min
Lab File: VX024817.D
Acq: 14 Oct 2021 16:37




Instrument :
MSVOA_X
ClientSampleId :
VX1014WBSD01


Tgt Ion: 94 Resp: 18840
Ion Ratio Lower Upper
94 100
96 92.8 74.8 112.2



#6
Chloroethane
Concen: 17.364 ug/l
RT: 1.684 min Scan# 98
Delta R.T. 0.006 min
Lab File: VX024817.D
Acq: 14 Oct 2021 16:37

Tgt Ion: 64 Resp: 16258
Ion Ratio Lower Upper
64 100
66 30.0 25.6 38.4









#7
Trichlorofluoromethane
Concen: 17.259 ug/l
RT: 1.886 min Scan# 131
Delta R.T. -0.000 min
Lab File: VX024817.D
Acq: 14 Oct 2021 16:37




Instrument :
MSVOA_X
ClientSampleId :
VX1014WBSD01


Tgt Ion:101 Resp: 44869
Ion Ratio Lower Upper
101 100
103 67.2 51.0 76.6



#8
Diethyl Ether
Concen: 18.556 ug/l
RT: 2.136 min Scan# 172
Delta R.T. -0.000 min
Lab File: VX024817.D
Acq: 14 Oct 2021 16:37

Tgt Ion: 74 Resp: 15840
Ion Ratio Lower Upper
74 100
45 100.5 47.6 142.9









#9
1,1,2-Trichlorotrifluoroethane
Concen: 17.408 ug/l
RT: 2.331 min Scan# 204
Delta R.T. -0.000 min
Lab File: VX024817.D
Acq: 14 Oct 2021 16:37




Instrument :
MSVOA_X
ClientSampleId :
VX1014WBSD01


Tgt Ion:101 Resp: 26613
Ion Ratio Lower Upper
101 100
85 45.8 35.9 53.9
151 81.3 64.4 96.6



#10
Methyl Iodide
Concen: 16.544 ug/l
RT: 2.453 min Scan# 224
Delta R.T. -0.000 min
Lab File: VX024817.D
Acq: 14 Oct 2021 16:37

Tgt Ion:142 Resp: 30711
Ion Ratio Lower Upper
142 100
127 48.6 38.9 58.3
141 14.7 11.7 17.5









#11

Tert butyl alcohol
Concen: 95.363 ug/l
RT: 2.977 min Scan# 310
Delta R.T. -0.000 min
Lab File: VX024817.D
Acq: 14 Oct 2021 16:37

Instrument :
MSVOA_X
ClientSampleId :
VX1014WBSD01




Tgt Ion: 59 Resp: 39115
Ion Ratio Lower Upper
59 100
57 10.3 8.5 12.7




#12

1,1-Dichloroethene
Concen: 17.231 ug/l
RT: 2.319 min Scan# 202
Delta R.T. -0.000 min
Lab File: VX024817.D
Acq: 14 Oct 2021 16:37

Tgt Ion: 96 Resp: 23802
Ion Ratio Lower Upper
96 100
61 167.2 134.1 201.1
98 63.6 51.5 77.3









#13
Acrolein
Concen: 103.099 ug/l
RT: 2.239 min Scan# 189
Delta R.T. -0.000 min
Lab File: VX024817.D
Acq: 14 Oct 2021 16:37




Instrument :
MSVOA_X
ClientSampleId :
VX1014WBSD01


Tgt Ion: 56 Resp: 17560
Ion Ratio Lower Upper
56 100
55 72.2 58.6 87.8



#14
Allyl chloride
Concen: 17.877 ug/l
RT: 2.666 min Scan# 259
Delta R.T. -0.000 min
Lab File: VX024817.D
Acq: 14 Oct 2021 16:37

Tgt Ion: 41 Resp: 44599
Ion Ratio Lower Upper
41 100
39 67.8 54.6 82.0
76 31.3 25.3 37.9









#15
Acrylonitrile
Concen: 96.950 ug/l
RT: 3.068 min Scan# 325
Delta R.T. -0.000 min
Lab File: VX024817.D
Acq: 14 Oct 2021 16:37




Instrument :
MSVOA_X
ClientSampleId :
VX1014WBSD01


Tgt Ion: 53 Resp: 88667
Ion Ratio Lower Upper
53 100
52 83.1 66.2 99.4
51 37.5 29.8 44.8



#16
Acetone
Concen: 73.732 ug/l
RT: 2.386 min Scan# 213
Delta R.T. -0.000 min
Lab File: VX024817.D
Acq: 14 Oct 2021 16:37

Tgt Ion: 43 Resp: 86150
Ion Ratio Lower Upper
43 100
58 29.0 24.0 36.0









#17
Carbon Disulfide
Concen: 15.804 ug/l
RT: 2.514 min Scan# 234
Delta R.T. -0.000 min
Lab File: VX024817.D
Acq: 14 Oct 2021 16:37




Instrument :
MSVOA_X
ClientSampleId :
VX1014WBSD01


Tgt Ion: 76 Resp: 57034
Ion Ratio Lower Upper
76 100
78 9.1 6.9 10.3



#18
Methyl Acetate
Concen: 19.814 ug/l
RT: 2.709 min Scan# 266
Delta R.T. -0.000 min
Lab File: VX024817.D
Acq: 14 Oct 2021 16:37

Tgt Ion: 43 Resp: 63647
Ion Ratio Lower Upper
43 100
74 23.9 19.0 28.6









#19
Methyl tert-butyl Ether
Concen: 18.420 ug/l
RT: 3.123 min Scan# 334
Delta R.T. -0.000 min
Lab File: VX024817.D
Acq: 14 Oct 2021 16:37




Instrument :
MSVOA_X
ClientSampleId :
VX1014WBSD01


Tgt Ion: 73 Resp: 93967
Ion Ratio Lower Upper
73 100
57 21.6 17.5 26.3



#20
Methylene Chloride
Concen: 18.168 ug/l
RT: 2.794 min Scan# 280
Delta R.T. -0.000 min
Lab File: VX024817.D
Acq: 14 Oct 2021 16:37

Tgt Ion: 84 Resp: 29983
Ion Ratio Lower Upper
84 100
49 117.6 98.2 147.2
51 39.3 30.2 45.4
86 65.0 52.2 78.2









#21
trans-1,2-Dichloroethene
Concen: 17.222 ug/l
RT: 3.093 min Scan# 329
Delta R.T. -0.000 min
Lab File: VX024817.D
Acq: 14 Oct 2021 16:37




Instrument :
MSVOA_X
ClientSampleId :
VX1014WBSD01


Tgt Ion: 96 Resp: 26546
Ion Ratio Lower Upper
96 100
61 144.4 117.9 176.9
98 61.9 51.1 76.7



#22
Diisopropyl ether
Concen: 18.472 ug/l
RT: 3.763 min Scan# 439
Delta R.T. -0.006 min
Lab File: VX024817.D
Acq: 14 Oct 2021 16:37

Tgt Ion: 45 Resp: 93023
Ion Ratio Lower Upper
45 100
43 88.1 58.3 87.5#
87 27.5 23.7 35.5
59 11.4 9.4 14.0









#23
 Vinyl Acetate
 Concen: 92.925 ug/l
 RT: 3.727 min Scan# 433
 Delta R.T. -0.006 min
 Lab File: VX024817.D
 Acq: 14 Oct 2021 16:37




Instrument :
 MSVOA_X
 ClientSampleId :
 VX1014WBSD01


Tgt Ion: 43 Resp: 359384
 Ion Ratio Lower Upper
 43 100
 86 10.3 8.2 12.4



#24
 1,1-Dichloroethane
 Concen: 17.708 ug/l
 RT: 3.611 min Scan# 414
 Delta R.T. -0.006 min
 Lab File: VX024817.D
 Acq: 14 Oct 2021 16:37

Tgt Ion: 63 Resp: 51303
 Ion Ratio Lower Upper
 63 100
 98 5.8 3.3 9.9
 100 3.5 2.1 6.2









#25
2-Butanone
Concen: 93.337 ug/l
RT: 4.568 min Scan# 571
Delta R.T. -0.000 min
Lab File: VX024817.D
Acq: 14 Oct 2021 16:37




Instrument :
MSVOA_X
ClientSampleId :
VX1014WBSD01


Tgt Ion: 43 Resp: 129264
Ion Ratio Lower Upper
43 100
72 25.4 21.2 31.8



#26
2,2-Dichloropropane
Concen: 16.495 ug/l
RT: 4.483 min Scan# 557
Delta R.T. -0.000 min
Lab File: VX024817.D
Acq: 14 Oct 2021 16:37

Tgt Ion: 77 Resp: 41556
Ion Ratio Lower Upper
77 100
97 22.5 11.2 33.5









#27
cis-1,2-Dichloroethene
Concen: 17.678 ug/l
RT: 4.495 min Scan# 559
Delta R.T. -0.000 min
Lab File: VX024817.D
Acq: 14 Oct 2021 16:37

Instrument :
MSVOA_X
ClientSampleId :
VX1014WBSD01

Tgt Ion: 96 Resp: 31240




Ion	Ratio	Lower	Upper
96	100		
61	149.6	0.0	302.2
98	65.2	0.0	131.4




#28
Bromochloromethane
Concen: 18.449 ug/l
RT: 4.903 min Scan# 626
Delta R.T. -0.006 min
Lab File: VX024817.D
Acq: 14 Oct 2021 16:37

Tgt Ion: 49 Resp: 22323

Ion	Ratio	Lower	Upper
49	100		
129	2.5	0.0	4.4
130	83.0	65.2	97.8





#29

Tetrahydrofuran

Concen: 94.438 ug/l

RT: 5.019 min Scan# 645

Delta R.T. 0.006 min





Lab File: VX024817.D

Acq: 14 Oct 2021 16:37

Tgt Ion: 42 Resp: 83111

Ion	Ratio	Lower	Upper
42	100		
72	45.0	36.1	54.1
71	41.1	33.7	50.5

Instrument :
MSVOA_X
ClientSampleId :
VX1014WBSD01



#30

Chloroform

Concen: 17.550 ug/l

RT: 5.098 min Scan# 658




Delta R.T. -0.000 min


Lab File: VX024817.D

Acq: 14 Oct 2021 16:37

Tgt Ion: 83 Resp: 54654

Ion	Ratio	Lower	Upper
83	100		
85	64.6	50.2	75.2









#31
Cyclohexane
Concen: 17.186 ug/l
RT: 5.476 min Scan# 720
Delta R.T. -0.000 min
Lab File: VX024817.D
Acq: 14 Oct 2021 16:37




Instrument :
MSVOA_X
ClientSampleId :
VX1014WBSD01


Tgt Ion: 56 Resp: 44553
Ion Ratio Lower Upper
56 100
69 32.1 24.2 36.2
84 86.9 71.5 107.3



#32
1,1,1-Trichloroethane
Concen: 17.395 ug/l
RT: 5.385 min Scan# 705
Delta R.T. -0.006 min
Lab File: VX024817.D
Acq: 14 Oct 2021 16:37

Tgt Ion: 97 Resp: 48361
Ion Ratio Lower Upper
97 100
99 63.8 50.9 76.3
61 42.8 35.0 52.4









#33

1,2-Dichloroethane-d4
Concen: 50.599 ug/l
RT: 5.964 min Scan# 800
Delta R.T. -0.000 min
Lab File: VX024817.D
Acq: 14 Oct 2021 16:37

Instrument :
MSVOA_X
ClientSampleId :
VX1014WBSD01




Tgt Ion: 65 Resp: 98464
Ion Ratio Lower Upper
65 100
67 51.2 0.0 101.0




#34

1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 6.769 min Scan# 932
Delta R.T. -0.000 min
Lab File: VX024817.D
Acq: 14 Oct 2021 16:37

Tgt Ion: 114 Resp: 247033
Ion Ratio Lower Upper
114 100
63 19.5 0.0 41.0
88 16.5 0.0 34.8









#35
Dibromofluoromethane
Concen: 50.452 ug/l
RT: 5.391 min Scan# 706
Delta R.T. -0.000 min
Lab File: VX024817.D
Acq: 14 Oct 2021 16:37




Instrument :
MSVOA_X
ClientSampleId :
VX1014WBSD01


Tgt Ion:	113	Resp:	74196
Ion	Ratio	Lower	Upper
113	100		
111	103.0	82.1	123.1
192	20.2	15.6	23.4



#36
1,1-Dichloropropene
Concen: 16.967 ug/l
RT: 5.696 min Scan# 756
Delta R.T. -0.006 min
Lab File: VX024817.D
Acq: 14 Oct 2021 16:37

Tgt Ion:	75	Resp:	37094
Ion	Ratio	Lower	Upper
75	100		
110	36.9	18.1	54.3
77	31.6	24.4	36.6









#37
Ethyl Acetate
Concen: 18.932 ug/l
RT: 4.727 min Scan# 597
Delta R.T. -0.000 min
Lab File: VX024817.D
Acq: 14 Oct 2021 16:37




Instrument :
MSVOA_X
ClientSampleId :
VX1014WBSD01


Tgt Ion: 43 Resp: 47684
Ion Ratio Lower Upper
43 100
61 13.7 11.4 17.0
70 10.9 8.6 12.8



#38
Carbon Tetrachloride
Concen: 17.245 ug/l
RT: 5.684 min Scan# 754
Delta R.T. -0.000 min
Lab File: VX024817.D
Acq: 14 Oct 2021 16:37

Tgt Ion: 117 Resp: 42241
Ion Ratio Lower Upper
117 100
119 91.2 75.6 113.4
121 29.0 23.7 35.5





#39

Methylcyclohexane

Concen: 17.404 ug/l

RT: 7.385 min Scan# 1033

Delta R.T. -0.000 min

Lab File: VX024817.D

Acq: 14 Oct 2021 16:37





Tgt Ion: 83 Resp: 47170

Ion Ratio Lower Upper

83 100

55 73.6 60.8 91.2

98 45.3 37.0 55.4



#40

Benzene

Concen: 17.985 ug/l

RT: 6.043 min Scan# 813

Delta R.T. -0.000 min

Lab File: VX024817.D




Acq: 14 Oct 2021 16:37


Tgt Ion: 78 Resp: 113007

Ion Ratio Lower Upper

78 100

77 24.1 18.3 27.5









#41
Methacrylonitrile
Concen: 19.338 ug/l
RT: 4.928 min Scan# 630
Delta R.T. -0.006 min
Lab File: VX024817.D
Acq: 14 Oct 2021 16:37




Instrument :
MSVOA_X
ClientSampleId :
VX1014WBSD01


Tgt Ion: 41 Resp: 26575
Ion Ratio Lower Upper
41 100
39 54.7 44.6 67.0
67 71.7 56.7 85.1
52 27.9 24.4 36.6



#42
1,2-Dichloroethane
Concen: 18.402 ug/l
RT: 6.092 min Scan# 821
Delta R.T. -0.000 min
Lab File: VX024817.D
Acq: 14 Oct 2021 16:37

Tgt Ion: 62 Resp: 47608
Ion Ratio Lower Upper
62 100
98 7.8 0.0 16.0









#43
Isopropyl Acetate
Concen: 18.775 ug/l
RT: 6.348 min Scan# 863
Delta R.T. -0.000 min
Lab File: VX024817.D
Acq: 14 Oct 2021 16:37




Instrument :
MSVOA_X
ClientSampleId :
VX1014WBSD01


Tgt Ion:	43	Resp:	73143
Ion	Ratio	Lower	Upper
43	100		
61	19.3	15.3	22.9
87	13.9	10.9	16.3



#44
Trichloroethene
Concen: 17.800 ug/l
RT: 7.129 min Scan# 991
Delta R.T. -0.000 min
Lab File: VX024817.D
Acq: 14 Oct 2021 16:37

Tgt Ion:	130	Resp:	31352
Ion	Ratio	Lower	Upper
130	100		
95	99.9	0.0	196.8









#45
1,2-Dichloropropane
Concen: 18.048 ug/l
RT: 7.433 min Scan# 1041
Delta R.T. -0.000 min
Lab File: VX024817.D
Acq: 14 Oct 2021 16:37




Instrument :
MSVOA_X
ClientSampleId :
VX1014WBSD01


Tgt Ion: 63 Resp: 29495
Ion Ratio Lower Upper
63 100
65 34.1 25.1 37.7



#46
Dibromomethane
Concen: 18.572 ug/l
RT: 7.586 min Scan# 1066
Delta R.T. -0.000 min
Lab File: VX024817.D
Acq: 14 Oct 2021 16:37

Tgt Ion: 93 Resp: 21967
Ion Ratio Lower Upper
93 100
95 83.0 66.6 100.0
174 100.3 79.2 118.8









#47

Bromodichloromethane
Concen: 18.098 ug/l
RT: 7.824 min Scan# 1105
Delta R.T. -0.006 min
Lab File: VX024817.D
Acq: 14 Oct 2021 16:37

Instrument :
MSVOA_X
ClientSampleId :
VX1014WBSD01




Tgt Ion:	83	Resp:	40473
Ion	Ratio	Lower	Upper
83	100		
85	63.4	53.5	80.3
127	8.2	7.7	11.5




#48

Methyl methacrylate
Concen: 19.012 ug/l
RT: 7.696 min Scan# 1084
Delta R.T. -0.000 min
Lab File: VX024817.D
Acq: 14 Oct 2021 16:37

Tgt Ion:	41	Resp:	36122
Ion	Ratio	Lower	Upper
41	100		
69	79.8	64.5	96.7
39	57.4	46.9	70.3





#49

1,4-Dioxane

Concen: 388.334 ug/l

RT: 7.665 min Scan# 1079

Delta R.T. -0.000 min

Lab File: VX024817.D

Acq: 14 Oct 2021 16:37

Tgt Ion: 88 Resp: 15901

Ion Ratio Lower Upper

88 100

43 33.8 27.0 40.6





58 71.3 55.2 82.8

Instrument :

MSVOA_X

ClientSampled :

VX1014WBSD01



#50

Toluene-d8

Concen: 50.619 ug/l

RT: 8.653 min Scan# 1241

Delta R.T. -0.000 min

Lab File: VX024817.D




Acq: 14 Oct 2021 16:37


Tgt Ion: 98 Resp: 275365

Ion Ratio Lower Upper

98 100

100 63.6 51.4 77.0









#51
4-Methyl-2-Pentanone
Concen: 96.433 ug/l
RT: 8.580 min Scan# 1229
Delta R.T. -0.000 min
Lab File: VX024817.D
Acq: 14 Oct 2021 16:37




Instrument :
MSVOA_X
ClientSampleId :
VX1014WBSD01


Tgt Ion: 43 Resp: 249363
Ion Ratio Lower Upper
43 100
58 38.9 30.9 46.3



#52
Toluene
Concen: 18.005 ug/l
RT: 8.720 min Scan# 1252
Delta R.T. -0.000 min
Lab File: VX024817.D
Acq: 14 Oct 2021 16:37

Tgt Ion: 92 Resp: 73206
Ion Ratio Lower Upper
92 100
91 174.5 138.6 208.0









#53
t-1,3-Dichloropropene
Concen: 17.620 ug/l
RT: 8.982 min Scan# 1295
Delta R.T. -0.000 min
Lab File: VX024817.D
Acq: 14 Oct 2021 16:37




Instrument :
MSVOA_X
ClientSampleId :
VX1014WBSD01


Tgt Ion: 75 Resp: 43785
Ion Ratio Lower Upper
75 100
77 31.5 25.6 38.4



#54
cis-1,3-Dichloropropene
Concen: 18.027 ug/l
RT: 8.372 min Scan# 1195
Delta R.T. -0.000 min
Lab File: VX024817.D
Acq: 14 Oct 2021 16:37

Tgt Ion: 75 Resp: 46932
Ion Ratio Lower Upper
75 100
77 30.1 23.8 35.8
39 46.6 38.8 58.2









#55
1,1,2-Trichloroethane
Concen: 18.807 ug/l
RT: 9.153 min Scan# 1323
Delta R.T. -0.000 min
Lab File: VX024817.D
Acq: 14 Oct 2021 16:37

Instrument :
MSVOA_X
ClientSampleId :
VX1014WBSD01

Tgt Ion: 97 Resp: 32310




Ion	Ratio	Lower	Upper
97	100		
83	83.6	69.2	103.8
85	56.2	43.3	64.9
99	63.2	49.1	73.7




#56
Ethyl methacrylate
Concen: 17.007 ug/l
RT: 9.122 min Scan# 1318
Delta R.T. 0.000 min
Lab File: VX024817.D
Acq: 14 Oct 2021 16:37

Tgt Ion: 69 Resp: 47077

Ion	Ratio	Lower	Upper
69	100		
41	68.0	55.3	82.9
39	35.5	30.2	45.4









#57

1,3-Dichloropropane
Concen: 19.188 ug/l
RT: 9.311 min Scan# 1349
Delta R.T. -0.000 min
Lab File: VX024817.D
Acq: 14 Oct 2021 16:37

Instrument :
MSVOA_X
ClientSampleId :
VX1014WBSD01




Tgt Ion: 76 Resp: 53741
Ion Ratio Lower Upper
76 100
78 31.6 25.8 38.8




#58

2-Chloroethyl Vinyl ether
Concen: 106.019 ug/l
RT: 8.244 min Scan# 1174
Delta R.T. -0.000 min
Lab File: VX024817.D
Acq: 14 Oct 2021 16:37

Tgt Ion: 63 Resp: 127299
Ion Ratio Lower Upper
63 100
106 28.5 22.8 34.2









#59
2-Hexanone
Concen: 86.540 ug/l
RT: 9.433 min Scan# 1369
Delta R.T. 0.000 min
Lab File: VX024817.D
Acq: 14 Oct 2021 16:37




Instrument :
MSVOA_X
ClientSampleId :
VX1014WBSD01


Tgt Ion: 43 Resp: 197065
Ion Ratio Lower Upper
43 100
58 53.6 26.6 79.7



#60
Dibromochloromethane
Concen: 16.673 ug/l
RT: 9.525 min Scan# 1384
Delta R.T. -0.000 min
Lab File: VX024817.D
Acq: 14 Oct 2021 16:37

Tgt Ion: 129 Resp: 30813
Ion Ratio Lower Upper
129 100
127 76.5 38.5 115.5









#61
1,2-Dibromoethane
Concen: 18.335 ug/l
RT: 9.610 min Scan# 1398
Delta R.T. -0.006 min
Lab File: VX024817.D
Acq: 14 Oct 2021 16:37




Instrument :
MSVOA_X
ClientSampleId :
VX1014WBSD01


Tgt Ion:107 Resp: 33857
Ion Ratio Lower Upper
107 100
109 95.9 74.4 111.6



#62
4-Bromofluorobenzene
Concen: 49.361 ug/l
RT: 11.085 min Scan# 1640
Delta R.T. -0.000 min
Lab File: VX024817.D
Acq: 14 Oct 2021 16:37

Tgt Ion: 95 Resp: 105610
Ion Ratio Lower Upper
95 100
174 79.3 0.0 158.0
176 76.4 0.0 153.6









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.055 min Scan# 1471
Delta R.T. -0.000 min
Lab File: VX024817.D
Acq: 14 Oct 2021 16:37




Instrument :
MSVOA_X
ClientSampleId :
VX1014WBSD01


Tgt Ion:117 Resp: 237385
Ion Ratio Lower Upper
117 100
82 56.4 44.6 67.0
119 30.9 24.8 37.2



#64
Tetrachloroethene
Concen: 17.602 ug/l
RT: 9.281 min Scan# 1344
Delta R.T. 0.006 min
Lab File: VX024817.D
Acq: 14 Oct 2021 16:37

Tgt Ion:164 Resp: 30580
Ion Ratio Lower Upper
164 100
166 127.4 103.2 154.8
129 88.7 81.0 121.4
131 84.3 76.6 115.0









#65
Chlorobenzene
Concen: 17.332 ug/l
RT: 10.079 min Scan# 1475
Delta R.T. -0.000 min
Lab File: VX024817.D
Acq: 14 Oct 2021 16:37




Instrument :
MSVOA_X
ClientSampleId :
VX1014WBSD01


Tgt Ion:112 Resp: 80555
Ion Ratio Lower Upper
112 100
114 32.9 25.8 38.6



#66
1,1,1,2-Tetrachloroethane
Concen: 17.848 ug/l
RT: 10.165 min Scan# 1489
Delta R.T. -0.000 min
Lab File: VX024817.D
Acq: 14 Oct 2021 16:37

Tgt Ion:131 Resp: 30777
Ion Ratio Lower Upper
131 100
133 92.2 48.1 144.4
119 62.3 32.0 96.0









#67
Ethyl Benzene
Concen: 17.185 ug/l
RT: 10.195 min Scan# 1494
Delta R.T. -0.000 min
Lab File: VX024817.D
Acq: 14 Oct 2021 16:37




Instrument :
MSVOA_X
ClientSampleId :
VX1014WBSD01


Tgt Ion: 91 Resp: 139208
Ion Ratio Lower Upper
91 100
106 31.5 24.7 37.1



#68
m/p-Xylenes
Concen: 35.112 ug/l
RT: 10.305 min Scan# 1512
Delta R.T. -0.000 min
Lab File: VX024817.D
Acq: 14 Oct 2021 16:37

Tgt Ion: 106 Resp: 109732
Ion Ratio Lower Upper
106 100
91 207.7 163.4 245.0









#69
o-Xylene
Concen: 17.816 ug/l
RT: 10.646 min Scan# 1568
Delta R.T. -0.000 min
Lab File: VX024817.D
Acq: 14 Oct 2021 16:37




Instrument :
MSVOA_X
ClientSampleId :
VX1014WBSD01


Tgt Ion:106 Resp: 54638
Ion Ratio Lower Upper
106 100
91 212.1 109.5 328.4



#70
Styrene
Concen: 16.059 ug/l
RT: 10.658 min Scan# 1570
Delta R.T. -0.001 min
Lab File: VX024817.D
Acq: 14 Oct 2021 16:37

Tgt Ion:104 Resp: 88802
Ion Ratio Lower Upper
104 100
78 51.9 41.3 61.9
103 55.7 43.5 65.3









#71
Bromoform
Concen: 16.431 ug/l
RT: 10.805 min Scan# 1594
Delta R.T. -0.000 min
Lab File: VX024817.D
Acq: 14 Oct 2021 16:37




Instrument :
MSVOA_X
ClientSampleId :
VX1014WBSD01


Tgt Ion:173 Resp: 22144
Ion Ratio Lower Upper
173 100
175 50.7 24.3 73.0
254 0.0 0.0 0.0



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.024 min Scan# 1794
Delta R.T. 0.000 min
Lab File: VX024817.D
Acq: 14 Oct 2021 16:37

Tgt Ion:152 Resp: 118203
Ion Ratio Lower Upper
152 100
115 69.5 43.3 129.8
150 164.3 0.0 351.2









#73
Isopropylbenzene
Concen: 18.213 ug/l
RT: 10.963 min Scan# 1620
Delta R.T. -0.000 min
Lab File: VX024817.D
Acq: 14 Oct 2021 16:37




Instrument :
MSVOA_X
ClientSampleId :
VX1014WBSD01


Tgt Ion:105 Resp: 142391
Ion Ratio Lower Upper
105 100
120 26.2 13.3 39.8



#74
N-ethyl acetate
Concen: 17.961 ug/l
RT: 10.847 min Scan# 1601
Delta R.T. -0.000 min
Lab File: VX024817.D
Acq: 14 Oct 2021 16:37

Tgt Ion: 43 Resp: 58825
Ion Ratio Lower Upper
43 100
70 41.5 33.7 50.5
55 26.1 19.7 29.5
61 23.8 19.0 28.4









#75
1,1,2,2-Tetrachloroethane
Concen: 18.836 ug/l
RT: 11.213 min Scan# 1661
Delta R.T. -0.000 min
Lab File: VX024817.D
Acq: 14 Oct 2021 16:37




Instrument :
MSVOA_X
ClientSampleId :
VX1014WBSD01


Tgt Ion: 83 Resp: 49945
Ion Ratio Lower Upper
83 100
131 10.8 5.3 16.1
85 65.7 32.3 96.9



#76
1,2,3-Trichloropropane
Concen: 23.023 ug/l
RT: 11.244 min Scan# 1666
Delta R.T. -0.000 min
Lab File: VX024817.D
Acq: 14 Oct 2021 16:37

Tgt Ion: 75 Resp: 56775
Ion Ratio Lower Upper
75 100
77 32.4 20.0 59.9









#77
Bromobenzene
Concen: 17.619 ug/l
RT: 11.201 min Scan# 1659
Delta R.T. -0.000 min
Lab File: VX024817.D
Acq: 14 Oct 2021 16:37




Instrument :
MSVOA_X
ClientSampleId :
VX1014WBSD01


Tgt Ion:156 Resp: 35394
Ion Ratio Lower Upper
156 100
77 160.8 78.3 234.8
158 100.6 48.9 146.6



#78
n-propylbenzene
Concen: 18.178 ug/l
RT: 11.305 min Scan# 1676
Delta R.T. -0.000 min
Lab File: VX024817.D
Acq: 14 Oct 2021 16:37

Tgt Ion: 91 Resp: 163732
Ion Ratio Lower Upper
91 100
120 23.8 11.7 35.0









#79
2-Chlorotoluene
Concen: 18.264 ug/l
RT: 11.366 min Scan# 1686
Delta R.T. -0.000 min
Lab File: VX024817.D
Acq: 14 Oct 2021 16:37




Instrument :
MSVOA_X
ClientSampleId :
VX1014WBSD01


Tgt Ion: 91 Resp: 101310
Ion Ratio Lower Upper
91 100
126 34.1 17.0 50.9



#80
1,3,5-Trimethylbenzene
Concen: 18.789 ug/l
RT: 11.457 min Scan# 1701
Delta R.T. -0.000 min
Lab File: VX024817.D
Acq: 14 Oct 2021 16:37

Tgt Ion: 105 Resp: 124607
Ion Ratio Lower Upper
105 100
120 47.3 23.8 71.5









#81
trans-1,4-Dichloro-2-butene
Concen: 16.177 ug/l
RT: 11.018 min Scan# 1629
Delta R.T. -0.006 min
Lab File: VX024817.D
Acq: 14 Oct 2021 16:37




Instrument :
MSVOA_X
ClientSampleId :
VX1014WBSD01


Tgt Ion: 75 Resp: 14149
Ion Ratio Lower Upper
75 100
53 109.5 81.9 122.9
89 48.4 38.2 57.2



#82
4-Chlorotoluene
Concen: 18.230 ug/l
RT: 11.457 min Scan# 1701
Delta R.T. -0.000 min
Lab File: VX024817.D
Acq: 14 Oct 2021 16:37

Tgt Ion: 91 Resp: 117519
Ion Ratio Lower Upper
91 100
126 29.8 14.9 44.5









#83
tert-Butylbenzene
Concen: 18.335 ug/l
RT: 11.719 min Scan# 1744
Delta R.T. -0.000 min
Lab File: VX024817.D
Acq: 14 Oct 2021 16:37




Instrument :
MSVOA_X
ClientSampleId :
VX1014WBSD01


Tgt Ion:119 Resp: 117841
Ion Ratio Lower Upper
119 100
91 59.9 30.3 90.9
134 23.1 11.4 34.1



#84
1,2,4-Trimethylbenzene
Concen: 18.365 ug/l
RT: 11.756 min Scan# 1750
Delta R.T. -0.000 min
Lab File: VX024817.D
Acq: 14 Oct 2021 16:37

Tgt Ion:105 Resp: 122779
Ion Ratio Lower Upper
105 100
120 44.6 22.1 66.3









#85
sec-Butylbenzene
Concen: 18.422 ug/l
RT: 11.896 min Scan# 1773
Delta R.T. -0.000 min
Lab File: VX024817.D
Acq: 14 Oct 2021 16:37




Instrument :
MSVOA_X
ClientSampleId :
VX1014WBSD01


Tgt Ion:105 Resp: 151566
Ion Ratio Lower Upper
105 100
134 20.0 10.2 30.4



#86
p-Isopropyltoluene
Concen: 18.595 ug/l
RT: 12.012 min Scan# 1792
Delta R.T. -0.000 min
Lab File: VX024817.D
Acq: 14 Oct 2021 16:37

Tgt Ion:119 Resp: 128394
Ion Ratio Lower Upper
119 100
134 25.8 13.2 39.5
91 24.8 12.4 37.0









#87
1,3-Dichlorobenzene
Concen: 18.093 ug/l
RT: 11.969 min Scan# 1785
Delta R.T. -0.000 min
Lab File: VX024817.D
Acq: 14 Oct 2021 16:37




Instrument :
MSVOA_X
ClientSampleId :
VX1014WBSD01


Tgt Ion:146 Resp: 68521
Ion Ratio Lower Upper
146 100
111 39.8 20.6 61.9
148 61.0 32.0 96.2



#88
1,4-Dichlorobenzene
Concen: 17.744 ug/l
RT: 12.042 min Scan# 1797
Delta R.T. -0.000 min
Lab File: VX024817.D
Acq: 14 Oct 2021 16:37

Tgt Ion:146 Resp: 68679
Ion Ratio Lower Upper
146 100
111 41.3 20.4 61.3
148 63.7 32.4 97.0









#89
n-Butylbenzene
Concen: 17.894 ug/l
RT: 12.335 min Scan# 1845
Delta R.T. -0.000 min
Lab File: VX024817.D
Acq: 14 Oct 2021 16:37

Instrument :
MSVOA_X
ClientSampleId :
VX1014WBSD01

Tgt Ion: 91 Resp: 108864




Ion	Ratio	Lower	Upper
91	100		
92	51.9	26.3	78.8
134	25.6	13.0	38.9




#90
Hexachloroethane
Concen: 16.741 ug/l
RT: 12.542 min Scan# 1879
Delta R.T. -0.000 min
Lab File: VX024817.D
Acq: 14 Oct 2021 16:37

Tgt Ion: 117 Resp: 18547

Ion	Ratio	Lower	Upper
117	100		
201	79.3	38.9	116.6









#91
1,2-Dichlorobenzene
Concen: 18.182 ug/l
RT: 12.341 min Scan# 1846
Delta R.T. 0.006 min
Lab File: VX024817.D
Acq: 14 Oct 2021 16:37




Instrument :
MSVOA_X
ClientSampleId :
VX1014WBSD01


Tgt Ion:146 Resp: 68064
Ion Ratio Lower Upper
146 100
111 41.6 21.2 63.6
148 62.7 32.1 96.5



#92
1,2-Dibromo-3-Chloropropane
Concen: 18.497 ug/l
RT: 12.945 min Scan# 1945
Delta R.T. -0.000 min
Lab File: VX024817.D
Acq: 14 Oct 2021 16:37

Tgt Ion: 75 Resp: 11221
Ion Ratio Lower Upper
75 100
155 85.5 43.1 129.3
157 105.7 53.6 160.9









#93
1,2,4-Trichlorobenzene
Concen: 18.000 ug/l
RT: 13.591 min Scan# 2051
Delta R.T. -0.000 min
Lab File: VX024817.D
Acq: 14 Oct 2021 16:37




Instrument :
MSVOA_X
ClientSampleId :
VX1014WBSD01


Tgt Ion:180 Resp: 41124
Ion Ratio Lower Upper
180 100
182 92.5 47.3 141.8
145 31.3 16.1 48.3



#94
Hexachlorobutadiene
Concen: 17.408 ug/l
RT: 13.731 min Scan# 2074
Delta R.T. 0.006 min
Lab File: VX024817.D
Acq: 14 Oct 2021 16:37

Tgt Ion:225 Resp: 19202
Ion Ratio Lower Upper
225 100
223 62.4 30.9 92.7
227 62.3 31.1 93.5









#95
Naphthalene
Concen: 16.495 ug/l
RT: 13.780 min Scan# 2082
Delta R.T. -0.000 min
Lab File: VX024817.D
Acq: 14 Oct 2021 16:37

Instrument :
MSVOA_X
ClientSampleId :
VX1014WBSD01

Tgt Ion:128 Resp: 136006
Ion Ratio Lower Upper
128 100
127 12.9 10.3 15.5
129 10.9 8.8 13.2



#96
1,2,3-Trichlorobenzene
Concen: 18.085 ug/l
RT: 13.963 min Scan# 2112
Delta R.T. -0.000 min
Lab File: VX024817.D
Acq: 14 Oct 2021 16:37

Tgt Ion:180 Resp: 41522
Ion Ratio Lower Upper
180 100
182 97.3 47.4 142.3
145 34.1 16.8 50.3

