

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101421\
 Data File : VX024811.D
 Acq On : 14 Oct 2021 14:01
 Operator : JC/MD
 Sample : VX1014WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX1014WBL01

Quant Time: Oct 14 14:28:18 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X101321W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 14 03:51:38 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	146945	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	238656	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	214988	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	95281	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	105329	58.32	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	116.64%
35) Dibromofluoromethane	5.397	113	79141	55.70	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	111.40%
50) Toluene-d8	8.653	98	286041	54.43	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	108.86%
62) 4-Bromofluorobenzene	11.085	95	98885	47.84	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	95.68%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101421\
Data File : VX024811.D
Acq On : 14 Oct 2021 14:01
Operator : JC/MD
Sample : VX1014WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX1014WBL01

Quant Time: Oct 14 14:28:18 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X101321W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 14 03:51:38 2021
Response via : Initial Calibration

