

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101421\
 Data File : VX024815.D
 Acq On : 14 Oct 2021 15:41
 Operator : JC/MD
 Sample : M4190-07RE
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TRIP-BLANKRE

Quant Time: Oct 14 16:03:58 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X101321W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 14 03:51:38 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	142082	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	229985	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	212488	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	92751	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	104045	59.58	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	119.16%#
35) Dibromofluoromethane	5.397	113	77087	56.30	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	112.60%
50) Toluene-d8	8.652	98	277658	54.82	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	109.64%
62) 4-Bromofluorobenzene	11.085	95	96049	48.22	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	96.44%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101421\
Data File : VX024815.D
Acq On : 14 Oct 2021 15:41
Operator : JC/MD
Sample : M4190-07RE
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TRIP-BLANKRE

Quant Time: Oct 14 16:03:58 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X101321W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 14 03:51:38 2021
Response via : Initial Calibration

