

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101422\
 Data File : VX032127.D
 Acq On : 14 Oct 2022 11:35
 Operator : JC/MD
 Sample : VX1014MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX1014MBL01

Quant Time: Oct 17 02:22:24 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 06 14:28:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	86939	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	129963	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	132481	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	79433	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	95107	52.970	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 105.940%	
35) Dibromofluoromethane	5.385	113	70435	47.879	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 95.760%	
50) Toluene-d8	8.647	98	222416	42.016	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 84.040%#	
62) 4-Bromofluorobenzene	11.079	95	91915	49.115	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 98.240%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101422\
Data File : VX032127.D
Acq On : 14 Oct 2022 11:35
Operator : JC/MD
Sample : VX1014MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX1014MBL01

Quant Time: Oct 17 02:22:24 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 06 14:28:46 2022
Response via : Initial Calibration

