

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101422\
 Data File : VX032138.D
 Acq On : 14 Oct 2022 16:40
 Operator : JC/MD
 Sample : N5151-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 256317-6

Quant Time: Oct 17 02:26:35 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 06 14:28:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	83684	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	124677	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	110331	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	76543	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	92824	53.710	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	107.420%
35) Dibromofluoromethane	5.385	113	69411	49.183	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.360%
50) Toluene-d8	8.647	98	219286	43.181	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	86.360%
62) 4-Bromofluorobenzene	11.079	95	94196	52.468	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	104.940%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101422\
Data File : VX032138.D
Acq On : 14 Oct 2022 16:40
Operator : JC/MD
Sample : N5151-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
256317-6

Quant Time: Oct 17 02:26:35 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 06 14:28:46 2022
Response via : Initial Calibration

