

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101422\
 Data File : VX032144.D
 Acq On : 14 Oct 2022 18:57
 Operator : JC/MD
 Sample : N5145-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-6-101322

Quant Time: Oct 17 02:28:25 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 06 14:28:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	79508	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	118855	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	100502	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	68500	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	88256	53.749	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	107.500%
35) Dibromofluoromethane	5.391	113	64971	48.292	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.580%
50) Toluene-d8	8.653	98	238324	49.229	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	98.460%
62) 4-Bromofluorobenzene	11.079	95	81602	47.680	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	95.360%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101422\
Data File : VX032144.D
Acq On : 14 Oct 2022 18:57
Operator : JC/MD
Sample : N5145-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-6-101322

Quant Time: Oct 17 02:28:25 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 06 14:28:46 2022
Response via : Initial Calibration

