

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101422\
 Data File : VX032141.D
 Acq On : 14 Oct 2022 17:47
 Operator : JC/MD
 Sample : N5146-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 STORAGE-BLANK-WATER-REF-5

Quant Time: Oct 17 02:27:32 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 06 14:28:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	84025	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	123313	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	130529	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	75671	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	90347	52.064	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	104.120%
35) Dibromofluoromethane	5.385	113	66841	47.886	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.780%
50) Toluene-d8	8.647	98	253355	50.442	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	100.880%
62) 4-Bromofluorobenzene	11.079	95	95302	53.671	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	107.340%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101422\
Data File : VX032141.D
Acq On : 14 Oct 2022 17:47
Operator : JC/MD
Sample : N5146-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-WATER-REF-5

Quant Time: Oct 17 02:27:32 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 06 14:28:46 2022
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.550 min Scan# 712
Delta R.T. -0.006 min
Lab File: VX032141.D
Acq: 14 Oct 2022 17:47




Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-WATER-REF-5


Tgt Ion:168 Resp: 84025
Ion Ratio Lower Upper
168 100
99 63.3 48.8 73.2



#33
1,2-Dichloroethane-d4
Concen: 52.064 ug/l
RT: 5.958 min Scan# 799
Delta R.T. 0.000 min
Lab File: VX032141.D
Acq: 14 Oct 2022 17:47

Tgt Ion: 65 Resp: 90347
Ion Ratio Lower Upper
65 100
67 51.0 0.0 102.4









#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 6.763 min Scan# 91
Delta R.T. -0.000 min
Lab File: VX032141.D
Acq: 14 Oct 2022 17:47




Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-WATER-REF-5


Tgt Ion:114 Resp: 123313
Ion Ratio Lower Upper
114 100
63 21.4 0.0 42.6
88 16.1 0.0 33.0



#35
Dibromofluoromethane
Concen: 47.886 ug/l
RT: 5.385 min Scan# 705
Delta R.T. 0.000 min
Lab File: VX032141.D
Acq: 14 Oct 2022 17:47

Tgt Ion:113 Resp: 66841
Ion Ratio Lower Upper
113 100
111 102.7 82.4 123.6
192 18.2 14.1 21.1









#50
Toluene-d8
Concen: 50.442 ug/l
RT: 8.647 min Scan# 11
Delta R.T. -0.006 min
Lab File: VX032141.D
Acq: 14 Oct 2022 17:47




Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-WATER-REF-5


Tgt Ion: 98 Resp: 253355
Ion Ratio Lower Upper
98 100
100 65.2 52.5 78.7



#62
4-Bromofluorobenzene
Concen: 53.671 ug/l
RT: 11.079 min Scan# 1639
Delta R.T. 0.000 min
Lab File: VX032141.D
Acq: 14 Oct 2022 17:47

Tgt Ion: 95 Resp: 95302
Ion Ratio Lower Upper
95 100
174 74.4 0.0 153.2
176 72.0 0.0 144.0









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.055 min Scan# 1471
Delta R.T. -0.000 min
Lab File: VX032141.D
Acq: 14 Oct 2022 17:47

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-WATER-REF-5

Tgt Ion:117 Resp: 130529
Ion Ratio Lower Upper
117 100
82 59.1 46.8 70.2
119 31.9 26.4 39.6



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.024 min Scan# 1794
Delta R.T. 0.000 min
Lab File: VX032141.D
Acq: 14 Oct 2022 17:47

Tgt Ion:152 Resp: 75671
Ion Ratio Lower Upper
152 100
115 63.9 43.4 130.1
150 157.2 0.0 346.8

