

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101424\
 Data File : VX043380.D
 Acq On : 14 Oct 2024 14:53
 Operator : JC/MD
 Sample : P4402-03 20X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 3 - Tote

Quant Time: Oct 15 01:46:38 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100124W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 02 16:50:57 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	86297	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.751	114	159358	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	133879	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	60553	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	69826	49.544	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	99.080%		
35) Dibromofluoromethane	5.379	113	51216	46.603	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	93.200%		
50) Toluene-d8	8.641	98	183956	48.301	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	96.600%		
62) 4-Bromofluorobenzene	11.079	95	67204	48.571	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery =	97.140%		
Target Compounds						
11) Tert butyl alcohol	3.008	59	9782	36.474	ug/l	99
16) Acetone	2.392	43	200631	340.884	ug/l	99
25) 2-Butanone	4.568	43	32961	40.844	ug/l	98
59) 2-Hexanone	9.452	43	1658	1.311	ug/l	74

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101424\
Data File : VX043380.D
Acq On : 14 Oct 2024 14:53
Operator : JC/MD
Sample : P4402-03 20X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
3 - Tote

Quant Time: Oct 15 01:46:38 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100124W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 02 16:50:57 2024
Response via : Initial Calibration

