

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101425\
 Data File : VX048171.D
 Acq On : 14 Oct 2025 16:56
 Operator : JC/MD
 Sample : Q3312-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 362Kingsland-001S

Quant Time: Oct 15 02:48:56 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 17 06:39:58 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	116081	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.751	114	223818	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	253264	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	128249	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	98540	53.331	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	106.660%
35) Dibromofluoromethane	5.373	113	82905	55.231	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	110.460%
50) Toluene-d8	8.635	98	275177	52.981	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	105.960%
62) 4-Bromofluorobenzene	11.067	95	117503	59.610	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	119.220%
Target Compounds						
						Qvalue
16) Acetone	2.374	43	7614	9.477	ug/l #	87
25) 2-Butanone	4.574	43	4715	4.705	ug/l #	90
51) 4-Methyl-2-Pentanone	8.561	43	5646	2.644	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101425\
Data File : VX048171.D
Acq On : 14 Oct 2025 16:56
Operator : JC/MD
Sample : Q3312-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
362Kingsland-001S

Quant Time: Oct 15 02:48:56 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 17 06:39:58 2025
Response via : Initial Calibration

