

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101425\
 Data File : VX048173.D
 Acq On : 14 Oct 2025 17:37
 Operator : JC/MD
 Sample : Q3325-08
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 STORAGE-BLANK-WATER-REF-5

Quant Time: Oct 15 02:49:37 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 17 06:39:58 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	111167	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.751	114	226396	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	230418	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	123442	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.940	65	96984	54.809	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 109.620%	
35) Dibromofluoromethane	5.373	113	78071	51.419	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 102.840%	
50) Toluene-d8	8.635	98	244684	46.573	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 93.140%	
62) 4-Bromofluorobenzene	11.061	95	116729	58.543	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 117.080%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101425\
Data File : VX048173.D
Acq On : 14 Oct 2025 17:37
Operator : JC/MD
Sample : Q3325-08
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-WATER-REF-5

Quant Time: Oct 15 02:49:37 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 17 06:39:58 2025
Response via : Initial Calibration

