

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101425\
 Data File : VX048174.D
 Acq On : 14 Oct 2025 17:57
 Operator : JC/MD
 Sample : Q3325-09
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 STORAGE-BLANK-WATER-REF-4

Quant Time: Oct 15 02:49:57 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 17 06:39:58 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	119484	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.751	114	246280	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	273767	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	130356	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.940	65	104185	54.780	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	109.560%
35) Dibromofluoromethane	5.379	113	84957	51.436	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	102.880%
50) Toluene-d8	8.634	98	283894	49.674	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.340%
62) 4-Bromofluorobenzene	11.061	95	122581	56.515	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	113.020%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101425\
Data File : VX048174.D
Acq On : 14 Oct 2025 17:57
Operator : JC/MD
Sample : Q3325-09
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-WATER-REF-4

Quant Time: Oct 15 02:49:57 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 17 06:39:58 2025
Response via : Initial Calibration

