

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101518\
 Data File : VX005353.D
 Acq On : 15 Oct 2018 19:28
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Oct 16 05:28:05 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X101218W.M
 Quant Title : SW846 8260
 QLast Update : Fri Oct 12 11:28:43 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	87	0.00
2 T	Dichlorodifluoromethane	0.391	0.418	-6.9	84	0.00
3 P	Chloromethane	0.566	0.542	4.2	84	0.00
4 C	Vinyl Chloride	0.574	0.551	4.0#	88	0.00
5 T	Bromomethane	0.356	0.312	12.4	84	0.00
6 T	Chloroethane	0.342	0.351	-2.6	88	0.00
7 T	Trichlorofluoromethane	0.766	0.779	-1.7	90	0.00
8 T	Diethyl Ether	0.327	0.320	2.1	89	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.442	0.454	-2.7	89	0.00
10 T	Methyl Iodide	0.440	0.541	-23.0#	94	0.00
11 T	Tert butyl alcohol	0.157	0.160	-1.9	93	0.00
12 CM	1,1-Dichloroethene	0.438	0.433	1.1#	88	0.00
13 T	Acrolein	0.105	0.075	28.6#	63	0.00
14 T	Allyl chloride	0.961	0.955	0.6	86	0.00
15 T	Acrylonitrile	0.355	0.346	2.5	92	0.00
16 T	Acetone	0.321	0.312	2.8	87	0.00
17 T	Carbon Disulfide	1.327	1.247	6.0	86	0.00
18 T	Methyl Acetate	0.913	0.944	-3.4	92	0.00
19 T	Methyl tert-butyl Ether	1.542	1.582	-2.6	90	0.00
20 T	Methylene Chloride	0.520	0.503	3.3	88	0.00
21 T	trans-1,2-Dichloroethene	0.485	0.468	3.5	90	0.00
22 T	Diisopropyl ether	1.615	1.528	5.4	80	0.00
23 T	Vinyl Acetate	1.411	1.355	4.0	85	0.00
24 P	1,1-Dichloroethane	0.926	0.877	5.3	88	0.00
25 T	2-Butanone	0.469	0.451	3.8	82	-0.01
26 T	2,2-Dichloropropane	0.692	0.642	7.2	87	0.00
27 T	cis-1,2-Dichloroethene	0.527	0.496	5.9	87	0.00
28 T	Bromochloromethane	0.422	0.414	1.9	87	0.00
29 T	Tetrahydrofuran	0.299	0.287	4.0	79	0.00
30 C	Chloroform	0.859	0.841	2.1#	87	0.00
31 T	Cyclohexane	0.763	0.736	3.5	81	0.00
32 T	1,1,1-Trichloroethane	0.747	0.734	1.7	88	0.00
33 S	1,2-Dichloroethane-d4	0.555	0.529	4.7	88	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	85	0.00
35 S	Dibromofluoromethane	0.331	0.310	6.3	85	0.00
36 T	1,1-Dichloropropene	0.484	0.443	8.5	81	0.00
37 T	Ethyl Acetate	0.623	0.574	7.9	83	-0.01
38 T	Carbon Tetrachloride	0.487	0.469	3.7	87	0.00
39 T	Methylcyclohexane	0.524	0.524	0.0	88	0.00
40 TM	Benzene	1.453	1.377	5.2	81	0.00
41 T	Methacrylonitrile	0.334	0.315	5.7	83	0.00
42 TM	1,2-Dichloroethane	0.513	0.483	5.8	85	0.00
43 T	Isopropyl Acetate	0.830	0.850	-2.4	89	0.00
44 TM	Trichloroethene	0.416	0.376	9.6	91	0.00
45 C	1,2-Dichloropropane	0.356	0.353	0.8#	90	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101518\
 Data File : VX005353.D
 Acq On : 15 Oct 2018 19:28
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Oct 16 05:28:05 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X101218W.M
 Quant Title : SW846 8260
 QLast Update : Fri Oct 12 11:28:43 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
46 T	Dibromomethane	0.235	0.235	0.0	88	0.00
47 T	Bromodichloromethane	0.435	0.458	-5.3	94	0.00
48 T	Methyl methacrylate	0.430	0.450	-4.7	88	0.00
49 T	1,4-Dioxane	0.010	0.010	0.0	91	0.00
50 S	Toluene-d8	1.137	1.182	-4.0	93	0.00
51 T	4-Methyl-2-Pentanone	0.574	0.610	-6.3	93	0.00
52 CM	Toluene	0.795	0.853	-7.3#	93	0.00
53 T	t-1,3-Dichloropropene	0.471	0.511	-8.5	93	0.00
54 T	cis-1,3-Dichloropropene	0.510	0.553	-8.4	94	0.00
55 T	1,1,2-Trichloroethane	0.337	0.353	-4.7	95	0.00
56 T	Ethyl methacrylate	0.491	0.544	-10.8	93	0.00
57 T	1,3-Dichloropropane	0.562	0.589	-4.8	95	0.00
58 T	2-Chloroethyl Vinyl ether	0.275	0.311	-13.1	95	0.00
59 T	2-Hexanone	0.462	0.486	-5.2	93	0.00
60 T	Dibromochloromethane	0.348	0.381	-9.5	95	0.00
61 T	1,2-Dibromoethane	0.353	0.371	-5.1	94	0.00
62 S	4-Bromofluorobenzene	0.427	0.442	-3.5	94	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	95	0.00
64 T	Tetrachloroethene	0.403	0.404	-0.2	97	0.00
65 PM	Chlorobenzene	1.064	1.023	3.9	95	0.00
66 T	1,1,1,2-Tetrachloroethane	0.379	0.379	0.0	95	0.00
67 C	Ethyl Benzene	1.729	1.732	-0.2#	94	0.00
68 T	m/p-Xylenes	0.671	0.678	-1.0	93	0.00
69 T	o-Xylene	0.647	0.651	-0.6	94	0.00
70 T	Styrene	1.065	1.096	-2.9	93	0.00
71 P	Bromoform	0.339	0.341	-0.6	98	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	94	0.00
73 T	Isopropylbenzene	2.972	3.131	-5.3	94	0.00
74 T	N-amyl acetate	1.434	1.456	-1.5	93	0.00
75 P	1,1,2,2-Tetrachloroethane	1.121	1.075	4.1	95	0.00
76 T	1,2,3-Trichloropropane	0.970	0.998	-2.9	98	0.00
77 T	Bromobenzene	0.850	0.857	-0.8	96	0.00
78 T	n-propylbenzene	3.420	3.595	-5.1	95	0.00
79 T	2-Chlorotoluene	2.095	2.112	-0.8	93	0.00
80 T	1,3,5-Trimethylbenzene	2.545	2.690	-5.7	94	0.00
81 T	trans-1,4-Dichloro-2-butene	0.319	0.321	-0.6	94	0.00
82 T	4-Chlorotoluene	2.483	2.510	-1.1	93	0.00
83 T	tert-Butylbenzene	2.578	2.679	-3.9	95	0.00
84 T	1,2,4-Trimethylbenzene	2.619	2.770	-5.8	94	0.00
85 T	sec-Butylbenzene	3.050	3.225	-5.7	95	0.00
86 T	p-Isopropyltoluene	2.757	2.926	-6.1	94	0.00
87 T	1,3-Dichlorobenzene	1.588	1.580	0.5	95	0.00
88 T	1,4-Dichlorobenzene	1.625	1.608	1.0	96	0.00
89 T	n-Butylbenzene	2.503	2.584	-3.2	94	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101518\
Data File : VX005353.D
Acq On : 15 Oct 2018 19:28
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Oct 16 05:28:05 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X101218W.M
Quant Title : SW846 8260
QLast Update : Fri Oct 12 11:28:43 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
90 T	Hexachloroethane	0.475	0.476	-0.2	95	0.00
91 T	1,2-Dichlorobenzene	1.621	1.600	1.3	95	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.274	0.269	1.8	96	0.00
93 T	1,2,4-Trichlorobenzene	1.189	1.231	-3.5	95	0.00
94 T	Hexachlorobutadiene	0.628	0.624	0.6	97	0.00
95 T	Naphthalene	3.448	3.786	-9.8	96	0.00
96 T	1,2,3-Trichlorobenzene	1.217	1.273	-4.6	97	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6