

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX101519\
 Data File : VX013080.D
 Acq On : 15 Oct 2019 16:04
 Operator : JC/SP
 Sample : K5336-04
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FA19-JMW6001

Quant Time: Oct 16 06:58:45 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X100819W.M
 Quant Title : SW846 8260
 QLast Update : Fri Oct 11 09:48:22 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	174670	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	286293	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	246038	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	96190	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	109855	50.32	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.64%	
35) Dibromofluoromethane	5.49	113	85592	52.09	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.18%	
50) Toluene-d8	8.71	98	336026	52.43	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.86%	
62) 4-Bromofluorobenzene	11.14	95	110699	44.45	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.90%	
Target Compounds						
					Qvalue	
4) Vinyl Chloride	1.40	62	3123	1.562	ug/l	89
16) Acetone	2.44	43	3039	2.641	ug/l	90
21) trans-1,2-Dichloroethene	3.15	96	1261	0.667	ug/l	80
27) cis-1,2-Dichloroethene	4.59	96	183894	84.326	ug/l	88
37) Ethyl Acetate	4.83	43	8073	2.868	ug/l #	81
44) Trichloroethene	7.21	130	89734	43.528	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX101519\
Data File : VX013080.D
Acq On : 15 Oct 2019 16:04
Operator : JC/SP
Sample : K5336-04
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FA19-JMW6001

Quant Time: Oct 16 06:58:45 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X100819W.M
Quant Title : SW846 8260
QLast Update : Fri Oct 11 09:48:22 2019
Response via : Initial Calibration

