

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX101520\
 Data File : VX018948.D
 Acq On : 15 Oct 2020 21:23
 Operator : JC/SP
 Sample : L4371-04 5X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 14B-4

Quant Time: Oct 16 05:09:16 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA X\METHOD\624X101520W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Thu Oct 15 15:32:58 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	4.99	128	43972	30.00	ug/l	0.01
28) 1,4-Difluorobenzene	6.83	114	252551	30.00	ug/l	0.00
57) Chlorobenzene-d5	10.10	117	216768	30.00	ug/l	0.00

System Monitoring Compounds

27) 1,2-Dichloroethane-d4	6.03	65	92155	32.85	ug/l	0.00
Spiked Amount	30.000	Range	50 - 169	Recovery	=	109.50%
60) 4-Bromofluorobenzene	11.12	95	97309	28.83	ug/l	0.00
Spiked Amount	30.000	Range	56 - 143	Recovery	=	96.10%
63) Toluene-d8	8.70	98	292680	30.91	ug/l	0.00
Spiked Amount	30.000	Range	66 - 137	Recovery	=	103.03%

Target Compounds

					Ovalue
15) Acetone	2.42	58	1144	2.830	ug/l 81
18) Methylene Chloride	2.83	84	2156	0.813	ug/l 92
25) Chloroform	5.18	83	338546	73.614	ug/l 99
41) Bromodichloromethane	7.88	83	34149	8.605	ug/l 99
53) Dibromochloromethane	9.56	129	2024	0.608	ug/l 96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX101520\
Data File : VX018948.D
Acq On : 15 Oct 2020 21:23
Operator : JC/SP
Sample : L4371-04 5X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
14B-4

Quant Time: Oct 16 05:09:16 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA X\METHOD\624X101520W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Thu Oct 15 15:32:58 2020
Response via : Initial Calibration

