

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX101520\
 Data File : VX018951.D
 Acq On : 15 Oct 2020 22:31
 Operator : JC/SP
 Sample : L4386-03 5X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 14B-3

Manual Integrations
 APPROVED

apatel
 10/16/2020 2:11:45 PM

Quant Time: Oct 16 05:16:13 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA X\METHOD\624X101520W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Thu Oct 15 15:32:58 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	4.98	128	43859	30.00	ug/l	0.00
28) 1,4-Difluorobenzene	6.83	114	246162	30.00	ug/l	0.00
57) Chlorobenzene-d5	10.10	117	212363	30.00	ug/l	0.00

System Monitoring Compounds

27) 1,2-Dichloroethane-d4	6.03	65	91795	32.81	ug/l	0.00
Spiked Amount 30.000	Range 50 - 169		Recovery	=	109.37%	
60) 4-Bromofluorobenzene	11.12	95	97431	29.47	ug/l	0.00
Spiked Amount 30.000	Range 56 - 143		Recovery	=	98.23%	
63) Toluene-d8	8.70	98	291279	31.40	ug/l	0.00
Spiked Amount 30.000	Range 66 - 137		Recovery	=	104.67%	

Target Compounds

					Ovalue
15) Acetone	2.42	58	1242m	3.080	ug/l
18) Methylene Chloride	2.83	84	2401	0.907	ug/l
25) Chloroform	5.18	83	342101	74.579	ug/l
41) Bromodichloromethane	7.88	83	32933	8.514	ug/l
53) Dibromochloromethane	9.57	129	2171	0.669	ug/l

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX101520\
Data File : VX018951.D
Acq On : 15 Oct 2020 22:31
Operator : JC/SP
Sample : L4386-03 5X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampled :
14B-3

Manual Integrations
APPROVED

apatel
10/16/2020 2:11:45 PM

Quant Time: Oct 16 05:16:13 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA X\METHOD\624X101520W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Thu Oct 15 15:32:58 2020
Response via : Initial Calibration

