

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101521\
 Data File : VX024823.D
 Acq On : 14 Oct 2021 19:43
 Operator : JC/MD
 Sample : M4103-10
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 C0AC6

Quant Time: Oct 14 23:05:09 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXML100721WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Oct 14 23:04:34 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.769	114	238413	50.00	ug/L	0.00
28) Chlorobenzene-d5	10.061	117	223613	50.00	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	12.024	152	101648	50.00	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.367	65	59603	40.38	ug/L	0.00
Spiked Amount 50.000	Range 60 - 135		Recovery =	80.76%		
7) Chloroethane-d5	1.666	69	50349	51.81	ug/L	0.00
Spiked Amount 50.000	Range 70 - 130		Recovery =	103.62%		
11) 1,1-Dichloroethene-d2	2.306	63	98677	34.05	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery =	68.10%		
21) 2-Butanone-d5	4.471	46	142119	117.11	ug/L	0.00
Spiked Amount 100.000	Range 40 - 130		Recovery =	117.11%		
24) Chloroform-d	5.068	84	150858	50.46	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	100.92%		
26) 1,2-Dichloroethane-d4	5.964	65	111568	52.77	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	105.54%		
32) Benzene-d6	5.982	84	298139	47.33	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	94.66%		
36) 1,2-Dichloropropane-d6	7.312	67	97240	49.36	ug/L	0.00
Spiked Amount 50.000	Range 70 - 120		Recovery =	98.72%		
41) Toluene-d8	8.653	98	266469	43.42	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery =	86.84%		
43) trans-1,3-Dichloroprop...	8.951	79	45800	46.92	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery =	93.84%		
47) 2-Hexanone-d5	9.390	63	100176	108.31	ug/L	0.00
Spiked Amount 100.000	Range 45 - 130		Recovery =	108.31%		
56) 1,1,2,2-Tetrachloroeth...	11.195	84	142629	52.56	ug/L	0.00
Spiked Amount 50.000	Range 65 - 120		Recovery =	105.12%		
66) 1,2-Dichlorobenzene-d4	12.323	152	105944	53.27	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery =	106.54%		
Target Compounds					Qvalue	
13) Acetone	2.386	43	2970	2.48	ug/L	91
25) Chloroform	5.099	83	8432	2.57	ug/L	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101521\
Data File : VX024823.D
Acq On : 14 Oct 2021 19:43
Operator : JC/MD
Sample : M4103-10
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
C0AC6

Quant Time: Oct 14 23:05:09 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXML100721WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Oct 14 23:04:34 2021
Response via : Initial Calibration

