

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101521\
 Data File : VX024828.D
 Acq On : 15 Oct 2021 13:43
 Operator : JC/MD
 Sample : M4103-13
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VHBLK002

Quant Time: Oct 18 01:30:01 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXML100721WMA.M
 Quant Title : VOC Analysis
 QLast Update : Mon Oct 18 01:29:23 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.769	114	242025	50.00	ug/L	0.00
28) Chlorobenzene-d5	10.055	117	230646	50.00	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	12.024	152	106010	50.00	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.368	65	56047	37.40	ug/L	0.00
Spiked Amount 50.000	Range 60 - 135		Recovery =	74.80%		
7) Chloroethane-d5	1.666	69	46866	47.51	ug/L	0.00
Spiked Amount 50.000	Range 70 - 130		Recovery =	95.02%		
11) 1,1-Dichloroethene-d2	2.307	63	96141	32.68	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery =	65.36%		
21) 2-Butanone-d5	4.459	46	149058	120.99	ug/L	0.00
Spiked Amount 100.000	Range 40 - 130		Recovery =	120.99%		
24) Chloroform-d	5.062	84	151024	49.76	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	99.52%		
26) 1,2-Dichloroethane-d4	5.958	65	114501	53.35	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	106.70%		
32) Benzene-d6	5.977	84	299054	46.03	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	92.06%		
36) 1,2-Dichloropropane-d6	7.312	67	97951	48.20	ug/L	0.00
Spiked Amount 50.000	Range 70 - 120		Recovery =	96.40%		
41) Toluene-d8	8.653	98	272107	42.99	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery =	85.98%		
43) trans-1,3-Dichloroprop...	8.952	79	47452	47.13	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery =	94.26%		
47) 2-Hexanone-d5	9.384	63	103914	108.92	ug/L	0.00
Spiked Amount 100.000	Range 45 - 130		Recovery =	108.92%		
56) 1,1,2,2-Tetrachloroeth...	11.195	84	145601	52.02	ug/L	0.00
Spiked Amount 50.000	Range 65 - 120		Recovery =	104.04%		
66) 1,2-Dichlorobenzene-d4	12.323	152	109310	52.70	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery =	105.40%		
Target Compounds					Qvalue	
25) Chloroform	5.093	83	11745	3.53	ug/L	83

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101521\
Data File : VX024828.D
Acq On : 15 Oct 2021 13:43
Operator : JC/MD
Sample : M4103-13
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VHBLK002

Quant Time: Oct 18 01:30:01 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXLM100721WMA.M
Quant Title : VOC Analysis
QLast Update : Mon Oct 18 01:29:23 2021
Response via : Initial Calibration

