

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\WX101524\
 Data File : VX043389.D
 Acq On : 15 Oct 2024 08:26
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Oct 16 07:12:05 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100124W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 02 16:50:57 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	50.000	50.000	0.0	101	0.00
2 T	Dichlorodifluoromethane	50.000	50.752	-1.5	100	0.00
3 P	Chloromethane	50.000	50.176	-0.4	101	0.00
4 C	Vinyl Chloride	50.000	50.740	-1.5#	102	0.00
5 T	Bromomethane	50.000	48.885	2.2	102	0.00
6 T	Chloroethane	50.000	50.576	-1.2	102	0.00
7 T	Trichlorofluoromethane	50.000	54.349	-8.7	114	0.00
8 T	Diethyl Ether	50.000	49.916	0.2	103	0.00
9 T	1,1,2-Trichlorotrifluoroeth	50.000	51.126	-2.3	104	0.00
10 T	Methyl Iodide	50.000	52.319	-4.6	104	0.00
11 T	Tert butyl alcohol	250.000	264.918	-6.0	105	0.00
12 CM	1,1-Dichloroethene	50.000	47.220	5.6#	97	0.00
13 T	Acrolein	250.000	210.426	15.8	85	0.00
14 T	Allyl chloride	50.000	51.142	-2.3	104	0.00
15 T	Acrylonitrile	250.000	248.859	0.5	100	0.00
16 T	Acetone	250.000	294.639	-17.9	129	0.00
17 T	Carbon Disulfide	50.000	45.156	9.7	94	0.00
18 T	Methyl Acetate	50.000	46.848	6.3	91	0.00
19 T	Methyl tert-butyl Ether	50.000	51.520	-3.0	104	0.00
20 T	Methylene Chloride	50.000	49.150	1.7	102	0.00
21 T	trans-1,2-Dichloroethene	50.000	49.961	0.1	102	0.00
22 T	Diisopropyl ether	50.000	51.877	-3.8	105	0.00
23 T	Vinyl Acetate	250.000	205.487	17.8	77	0.00
24 P	1,1-Dichloroethane	50.000	50.602	-1.2	102	0.00
25 T	2-Butanone	250.000	264.591	-5.8	106	0.00
26 T	2,2-Dichloropropane	50.000	52.235	-4.5	107	0.00
27 T	cis-1,2-Dichloroethene	50.000	50.216	-0.4	103	0.00
28 T	Bromochloromethane	50.000	45.104	9.8	95	0.00
29 T	Tetrahydrofuran	250.000	234.414	6.2	97	0.00
30 C	Chloroform	50.000	50.405	-0.8#	104	0.00
31 T	Cyclohexane	50.000	49.445	1.1	103	0.00
32 T	1,1,1-Trichloroethane	50.000	49.404	1.2	100	0.00
33 S	1,2-Dichloroethane-d4	50.000	48.169	3.7	98	0.00
34 I	1,4-Difluorobenzene	50.000	50.000	0.0	100	0.00
35 S	Dibromofluoromethane	50.000	48.319	3.4	98	0.00
36 T	1,1-Dichloropropene	50.000	49.903	0.2	100	0.00
37 T	Ethyl Acetate	50.000	49.504	1.0	100	0.00
38 T	Carbon Tetrachloride	50.000	51.031	-2.1	100	0.00
39 T	Methylcyclohexane	50.000	51.633	-3.3	103	0.00
40 TM	Benzene	50.000	50.466	-0.9	102	0.00
41 T	Methacrylonitrile	50.000	51.887	-3.8	98	0.00
42 TM	1,2-Dichloroethane	50.000	51.838	-3.7	103	0.00
43 T	Isopropyl Acetate	50.000	51.075	-2.2	99	0.00
44 TM	Trichloroethene	50.000	49.249	1.5	101	0.00
45 C	1,2-Dichloropropane	50.000	50.737	-1.5#	101	0.00
46 T	Dibromomethane	50.000	51.837	-3.7	103	0.00
47 T	Bromodichloromethane	50.000	52.497	-5.0	102	0.00
48 T	Methyl methacrylate	50.000	51.241	-2.5	100	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\WX101524\
 Data File : VX043389.D
 Acq On : 15 Oct 2024 08:26
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Oct 16 07:12:05 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100124W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 02 16:50:57 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	1000.000	947.631	5.2	92	0.00
50 S	Toluene-d8	50.000	47.186	5.6	93	0.00
51 T	4-Methyl-2-Pentanone	250.000	248.262	0.7	96	0.00
52 CM	Toluene	50.000	51.698	-3.4#	101	0.00
53 T	t-1,3-Dichloropropene	50.000	54.277	-8.6	102	0.00
54 T	cis-1,3-Dichloropropene	50.000	52.311	-4.6	102	0.00
55 T	1,1,2-Trichloroethane	50.000	51.520	-3.0	100	0.00
56 T	Ethyl methacrylate	50.000	53.243	-6.5	99	0.00
57 T	1,3-Dichloropropane	50.000	51.159	-2.3	99	0.00
58 T	2-Chloroethyl Vinyl ether	250.000	262.990	-5.2	96	0.00
59 T	2-Hexanone	250.000	255.333	-2.1	98	0.00
60 T	Dibromochloromethane	50.000	52.005	-4.0	98	0.00
61 T	1,2-Dibromoethane	50.000	51.230	-2.5	99	0.00
62 S	4-Bromofluorobenzene	50.000	49.393	1.2	95	0.00
63 I	Chlorobenzene-d5	50.000	50.000	0.0	98	0.00
64 T	Tetrachloroethene	50.000	50.605	-1.2	101	0.00
65 PM	Chlorobenzene	50.000	49.798	0.4	100	0.00
66 T	1,1,1,2-Tetrachloroethane	50.000	51.015	-2.0	98	0.00
67 C	Ethyl Benzene	50.000	50.606	-1.2#	98	0.00
68 T	m/p-Xylenes	100.000	100.216	-0.2	98	0.00
69 T	o-Xylene	50.000	49.807	0.4	99	0.00
70 T	Styrene	50.000	51.132	-2.3	98	0.00
71 P	Bromoform	50.000	49.672	0.7	93	0.00
72 I	1,4-Dichlorobenzene-d4	50.000	50.000	0.0	98	0.00
73 T	Isopropylbenzene	50.000	49.843	0.3	97	0.00
74 T	N-amyl acetate	50.000	48.871	2.3	96	0.00
75 P	1,1,2,2-Tetrachloroethane	50.000	49.497	1.0	98	0.00
76 T	1,2,3-Trichloropropane	50.000	48.894	2.2	99	0.00
77 T	Bromobenzene	50.000	48.723	2.6	97	0.00
78 T	n-propylbenzene	50.000	52.079	-4.2	99	0.00
79 T	2-Chlorotoluene	50.000	49.252	1.5	99	0.00
80 T	1,3,5-Trimethylbenzene	50.000	50.386	-0.8	99	0.00
81 T	trans-1,4-Dichloro-2-butene	50.000	49.948	0.1	100	0.00
82 T	4-Chlorotoluene	50.000	50.142	-0.3	100	0.00
83 T	tert-Butylbenzene	50.000	49.282	1.4	97	0.00
84 T	1,2,4-Trimethylbenzene	50.000	50.379	-0.8	98	0.00
85 T	sec-Butylbenzene	50.000	50.808	-1.6	99	0.00
86 T	p-Isopropyltoluene	50.000	51.272	-2.5	99	0.00
87 T	1,3-Dichlorobenzene	50.000	50.706	-1.4	101	0.00
88 T	1,4-Dichlorobenzene	50.000	48.687	2.6	100	0.00
89 T	n-Butylbenzene	50.000	53.843	-7.7	102	0.00
90 T	Hexachloroethane	50.000	50.207	-0.4	95	0.00
91 T	1,2-Dichlorobenzene	50.000	49.051	1.9	99	0.00
92 T	1,2-Dibromo-3-Chloropropane	50.000	46.910	6.2	91	0.00
93 T	1,2,4-Trichlorobenzene	50.000	53.047	-6.1	102	0.00
94 T	Hexachlorobutadiene	50.000	53.920	-7.8	102	0.00
95 T	Naphthalene	50.000	49.020	2.0	97	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101524\
Data File : VX043389.D
Acq On : 15 Oct 2024 08:26
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Oct 16 07:12:05 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100124W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 02 16:50:57 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	50.000	50.559	-1.1	99	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6