

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101525\
 Data File : VX048182.D
 Acq On : 15 Oct 2025 10:26
 Operator : JC/MD
 Sample : VX1015MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX1015MBL01

Quant Time: Oct 16 04:01:59 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 17 06:39:58 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.537	168	136060	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.739	114	280699	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	278520	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	134008	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.934	65	117474	54.242	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	108.480%
35) Dibromofluoromethane	5.367	113	98372	52.255	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	104.520%
50) Toluene-d8	8.628	98	332682	51.073	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	102.140%
62) 4-Bromofluorobenzene	11.061	95	138838	56.161	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	112.320%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101525\
Data File : VX048182.D
Acq On : 15 Oct 2025 10:26
Operator : JC/MD
Sample : VX1015MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX1015MBL01

Quant Time: Oct 16 04:01:59 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 17 06:39:58 2025
Response via : Initial Calibration

