

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101525\
 Data File : VX048183.D
 Acq On : 15 Oct 2025 10:47
 Operator : JC/MD
 Sample : VX1015WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX1015WBL01

Quant Time: Oct 16 04:02:25 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 17 06:39:58 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	116070	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.745	114	256182	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	266771	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	133528	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	97588	52.820	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	105.640%
35) Dibromofluoromethane	5.373	113	79378	46.201	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.400%
50) Toluene-d8	8.634	98	282591	47.535	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	95.060%
62) 4-Bromofluorobenzene	11.067	95	124695	55.267	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	110.540%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101525\
Data File : VX048183.D
Acq On : 15 Oct 2025 10:47
Operator : JC/MD
Sample : VX1015WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX1015WBL01

Quant Time: Oct 16 04:02:25 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 17 06:39:58 2025
Response via : Initial Calibration

