

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101525\
 Data File : VX048187.D
 Acq On : 15 Oct 2025 13:54
 Operator : JC/MD
 Sample : Q3343-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SOIL-ROLLOFF-WC-1

Quant Time: Oct 16 04:05:01 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 17 06:39:58 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	137115	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.745	114	275742	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	297788	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	142742	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.940	65	108331	49.636	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	99.280%
35) Dibromofluoromethane	5.373	113	94978	51.360	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	102.720%
50) Toluene-d8	8.635	98	326202	50.978	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	101.960%
62) 4-Bromofluorobenzene	11.061	95	137927	56.795	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	113.600%
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	16105	16.971	ug/l	98
18) Methyl Acetate	2.703	43	7043	3.335	ug/l	97
20) Methylene Chloride	2.794	84	4486	2.233	ug/l	93
25) 2-Butanone	4.556	43	6519	5.507	ug/l	97
43) Isopropyl Acetate	6.330	43	8709	1.835	ug/l #	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101525\
Data File : VX048187.D
Acq On : 15 Oct 2025 13:54
Operator : JC/MD
Sample : Q3343-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SOIL-ROLLOFF-WC-1

Quant Time: Oct 16 04:05:01 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 17 06:39:58 2025
Response via : Initial Calibration

